

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051723\
 Data File : VX035714.D
 Acq On : 17 May 2023 13:38
 Operator : JC/MD
 Sample : 02812-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WG-0516-B3-GW

Quant Time: May 18 04:40:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	174071	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	327726	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	299321	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	134236	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	151351	49.840	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.680%
35) Dibromofluoromethane	5.385	113	106584	48.224	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.440%
50) Toluene-d8	8.647	98	397649	48.702	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.400%
62) 4-Bromofluorobenzene	11.079	95	161383	49.089	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.180%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	7338	5.622	ug/l	93
17) Carbon Disulfide	2.514	76	1180	0.252	ug/l	# 75
25) 2-Butanone	4.581	43	1663	0.920	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051723\
Data File : VX035714.D
Acq On : 17 May 2023 13:38
Operator : JC/MD
Sample : 02812-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WG-0516-B3-GW

Quant Time: May 18 04:40:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

