

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051723\
 Data File : VX035717.D
 Acq On : 17 May 2023 14:50
 Operator : JC/MD
 Sample : 02723-06ME
 Misc : 4.89g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CB-14-G1-SO-30.5-050923ME

Quant Time: May 18 04:41:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	192931	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	359974	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	328352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	147448	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	163676	48.630	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.260%
35) Dibromofluoromethane	5.385	113	106848	44.013	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.020%
50) Toluene-d8	8.647	98	431531	48.117	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.240%
62) 4-Bromofluorobenzene	11.085	95	176126	48.774	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.540%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.306	101	5018	2.220	ug/l	94
44) Trichloroethene	7.123	130	18970	6.979	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051723\
Data File : VX035717.D
Acq On : 17 May 2023 14:50
Operator : JC/MD
Sample : 02723-06ME
Misc : 4.89g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CB-14-G1-SO-30.5-050923ME

Quant Time: May 18 04:41:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

