

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051723\
 Data File : VX035719.D
 Acq On : 17 May 2023 15:37
 Operator : JC/MD
 Sample : 02723-08ME
 Misc : 6.06g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CB-14-G1-SO-35.5-050923ME

Quant Time: May 18 04:42:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	182960	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	341290	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	308741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135631	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	156360	48.988	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.980%		
35) Dibromofluoromethane	5.385	113	101376	44.045	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	88.080%		
50) Toluene-d8	8.647	98	412697	48.536	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.080%		
62) 4-Bromofluorobenzene	11.085	95	165579	48.364	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.720%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.306	101	4708	2.196	ug/l	94
44) Trichloroethene	7.123	130	15031	5.833	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051723\
Data File : VX035719.D
Acq On : 17 May 2023 15:37
Operator : JC/MD
Sample : 02723-08ME
Misc : 6.06g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CB-14-G1-SO-35.5-050923ME

Quant Time: May 18 04:42:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

