

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051723\
 Data File : VX035723.D
 Acq On : 17 May 2023 17:10
 Operator : JC/MD
 Sample : 02723-10ME
 Misc : 5.39g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CB-14-G1-SO-49.5-050923ME

Quant Time: May 18 04:42:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	192324	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	355409	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	326133	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	147087	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	163229	48.650	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.300%
35) Dibromofluoromethane	5.385	113	104924	43.775	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.560%
50) Toluene-d8	8.647	98	431381	48.718	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.440%
62) 4-Bromofluorobenzene	11.085	95	175430	49.206	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.420%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.306	101	85427	37.908	ug/l	95
27) cis-1,2-Dichloroethene	4.483	96	5111	1.812	ug/l	92
44) Trichloroethene	7.123	130	286843	106.884	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051723\
Data File : VX035723.D
Acq On : 17 May 2023 17:10
Operator : JC/MD
Sample : 02723-10ME
Misc : 5.39g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CB-14-G1-SO-49.5-050923ME

Quant Time: May 18 04:42:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

