



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051723\
 Data File : VX035703.D
 Acq On : 17 May 2023 08:31
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Title : SW846 8260
 Last Update : Thu May 11 13:57:54 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result
50	95	15	40	20.6	11583	PASS
75	95	30	60	55.5	31205	PASS
95	95	100	100	100.0	56195	PASS
96	95	5	9	6.8	3843	PASS
173	174	0.00	2	1.2	445	PASS
174	95	50	100	63.4	35651	PASS
175	174	5	9	7.3	2595	PASS
176	174	95	101	98.0	34949	PASS
177	176	5	9	6.7	2338	PASS