

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051820\
 Data File : VX016311.D
 Acq On : 18 May 2020 20:20
 Operator : JC/SP
 Sample : L2672-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP021MW038-200514

Quant Time: May 19 12:49:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	259848	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	431403	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	421206	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	215423	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	152132	44.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.98%	
35) Dibromofluoromethane	5.47	113	123738	45.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.54%	
50) Toluene-d8	8.70	98	529417	50.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.54%	
62) 4-Bromofluorobenzene	11.12	95	213484	53.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.60%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.31	50	1324	0.415	ug/l	91
4) Vinyl Chloride	1.39	62	807	0.237	ug/l	94
16) Acetone	2.42	43	4448	2.915	ug/l	100
24) 1,1-Dichloroethane	3.68	63	2551	0.488	ug/l #	95
27) cis-1,2-Dichloroethene	4.57	96	22685	6.894	ug/l	83
44) Trichloroethene	7.19	130	1555	0.352	ug/l	99
65) Chlorobenzene	10.12	112	22954	2.603	ug/l	99
91) 1,2-Dichlorobenzene	12.38	146	8023	1.147	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051820\
Data File : VX016311.D
Acq On : 18 May 2020 20:20
Operator : JC/SP
Sample : L2672-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP021MW038-200514

Quant Time: May 19 12:49:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

