

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX051820\
 Data File : VX016320.D
 Acq On : 18 May 2020 23:44
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 19 03:30:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.529	0.530	-0.2	78	0.00
3 P	Chloromethane	0.615	0.565	8.1	74	0.00
4 C	Vinyl Chloride	0.656	0.709	-8.1#	88	0.00
5 T	Bromomethane	0.331	0.258	22.1	64	0.00
6 T	Chloroethane	0.399	0.388	2.8	77	0.00
7 T	Trichlorofluoromethane	0.861	0.844	2.0	78	0.00
8 T	Diethyl Ether	0.363	0.356	1.9	80	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.497	0.478	3.8	77	0.00
10 T	Methyl Iodide	0.667	0.337	49.5#	39#	0.00
11 T	Tert butyl alcohol	0.166	0.156	6.0	75	0.00
12 CM	1,1-Dichloroethene	0.520	0.505	2.9#	79	0.00
13 T	Acrolein	0.104	0.113	-8.7	89	0.00
14 T	Allyl chloride	0.953	0.880	7.7	75	0.00
15 T	Acrylonitrile	0.345	0.347	-0.6	81	0.00
16 T	Acetone	0.294	0.265	9.9	73	0.00
17 T	Carbon Disulfide	1.769	1.362	23.0	70	0.00
18 T	Methyl Acetate	0.722	0.765	-6.0	86	0.00
19 T	Methyl tert-butyl Ether	1.792	1.893	-5.6	84	0.00
20 T	Methylene Chloride	0.605	0.593	2.0	82	0.00
21 T	trans-1,2-Dichloroethene	0.579	0.563	2.8	82	0.00
22 T	Diisopropyl ether	1.798	1.800	-0.1	81	0.00
23 T	Vinyl Acetate	1.582	1.562	1.3	78	0.00
24 P	1,1-Dichloroethane	1.006	1.025	-1.9	82	0.00
25 T	2-Butanone	0.473	0.467	1.3	78	0.00
26 T	2,2-Dichloropropane	0.918	0.755	17.8	67	0.00
27 T	cis-1,2-Dichloroethene	0.633	0.683	-7.9	88	0.00
28 T	Bromochloromethane	0.459	0.430	6.3	78	0.00
29 T	Tetrahydrofuran	0.313	0.306	2.2	78	0.00
30 C	Chloroform	1.001	1.045	-4.4#	84	0.00
31 T	Cyclohexane	0.914	0.837	8.4	73	0.00
32 T	1,1,1-Trichloroethane	0.910	0.932	-2.4	82	0.00
33 S	1,2-Dichloroethane-d4	0.658	0.615	6.5	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
35 S	Dibromofluoromethane	0.317	0.309	2.5	85	0.00
36 T	1,1-Dichloropropene	0.492	0.469	4.7	81	0.00
37 T	Ethyl Acetate	0.570	0.537	5.8	77	0.00
38 T	Carbon Tetrachloride	0.500	0.486	2.8	81	0.00
39 T	Methylcyclohexane	0.593	0.450	24.1	63	0.00
40 TM	Benzene	1.440	1.445	-0.3	84	0.00
41 T	Methacrylonitrile	0.308	0.292	5.2	77	0.00
42 TM	1,2-Dichloroethane	0.519	0.491	5.4	80	0.00
43 T	Isopropyl Acetate	0.910	0.869	4.5	79	0.00
44 TM	Trichloroethene	0.512	0.444	13.3	75	0.00
45 C	1,2-Dichloropropane	0.364	0.371	-1.9#	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX051820\
 Data File : VX016320.D
 Acq On : 18 May 2020 23:44
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 19 03:30:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.248	0.246	0.8	83	0.00
47 T	Bromodichloromethane	0.504	0.516	-2.4	83	0.00
48 T	Methyl methacrylate	0.431	0.412	4.4	78	0.00
49 T	1,4-Dioxane	0.010	0.009	10.0	73	0.00
50 S	Toluene-d8	1.221	1.188	2.7	84	0.00
51 T	4-Methyl-2-Pentanone	0.557	0.548	1.6	79	0.00
52 CM	Toluene	0.898	0.924	-2.9#	85	0.00
53 T	t-1,3-Dichloropropene	0.601	0.583	3.0	81	0.00
54 T	cis-1,3-Dichloropropene	0.628	0.618	1.6	82	0.00
55 T	1,1,2-Trichloroethane	0.353	0.373	-5.7	86	0.00
56 T	Ethyl methacrylate	0.576	0.604	-4.9	84	0.00
57 T	1,3-Dichloropropane	0.616	0.635	-3.1	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.307	0.291	5.2	77	0.00
59 T	2-Hexanone	0.437	0.427	2.3	79	0.00
60 T	Dibromochloromethane	0.389	0.427	-9.8	88	0.00
61 T	1,2-Dibromoethane	0.383	0.394	-2.9	83	0.00
62 S	4-Bromofluorobenzene	0.464	0.475	-2.4	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64 T	Tetrachloroethene	0.614	0.512	16.6	72	0.00
65 PM	Chlorobenzene	1.047	1.092	-4.3	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.387	0.406	-4.9	88	0.00
67 C	Ethyl Benzene	1.868	1.832	1.9#	82	0.00
68 T	m/p-Xylenes	0.700	0.705	-0.7	83	0.00
69 T	o-Xylene	0.673	0.687	-2.1	84	0.00
70 T	Styrene	1.141	1.198	-5.0	86	0.00
71 P	Bromoform	0.315	0.345	-9.5	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	3.637	3.554	2.3	81	0.00
74 T	N-amyl acetate	1.696	1.643	3.1	80	0.00
75 P	1,1,2,2-Tetrachloroethane	0.641	1.013	-58.0#	118	0.00
76 T	1,2,3-Trichloropropane	1.135	1.131	0.4	82	0.00
77 T	Bromobenzene	0.939	0.938	0.1	85	0.00
78 T	n-propylbenzene	4.231	3.937	6.9	78	0.00
79 T	2-Chlorotoluene	2.497	2.462	1.4	83	0.00
80 T	1,3,5-Trimethylbenzene	3.071	2.926	4.7	79	0.00
81 T	trans-1,4-Dichloro-2-butene	0.476	0.450	5.5	77	0.00
82 T	4-Chlorotoluene	2.949	2.888	2.1	82	0.00
83 T	tert-Butylbenzene	2.646	2.526	4.5	80	0.00
84 T	1,2,4-Trimethylbenzene	3.104	3.005	3.2	80	0.00
85 T	sec-Butylbenzene	3.573	2.955	17.3	69	0.00
86 T	p-Isopropyltoluene	3.317	2.778	16.2	69	0.00
87 T	1,3-Dichlorobenzene	1.685	1.600	5.0	81	0.00
88 T	1,4-Dichlorobenzene	1.711	1.682	1.7	84	0.00
89 T	n-Butylbenzene	3.030	2.257	25.5#	63	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX051820\
Data File : VX016320.D
Acq On : 18 May 2020 23:44
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 19 03:30:52 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.570	0.497	12.8	71	0.00
91 T	1,2-Dichlorobenzene	1.623	1.638	-0.9	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.299	0.291	2.7	83	0.00
93 T	1,2,4-Trichlorobenzene	1.202	0.994	17.3	73	0.00
94 T	Hexachlorobutadiene	0.556	0.332	40.3#	53	0.00
95 T	Naphthalene	3.891	3.780	2.9	82	0.00
96 T	1,2,3-Trichlorobenzene	1.174	1.004	14.5	73	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6