

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
 Data File : VX022038.D
 Acq On : 17 May 2021 17:49
 Operator : JC/MD
 Sample : VX0518WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0518WBL01

Quant Time: May 18 02:20:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	71331	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	159198	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171123	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75921	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	58985	50.25	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.50%	
35) Dibromofluoromethane	5.397	113	51085	46.82	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.64%	
50) Toluene-d8	8.653	98	205751	49.42	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.84%	
62) 4-Bromofluorobenzene	11.085	95	90094	54.83	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.66%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
Data File : VX022038.D
Acq On : 17 May 2021 17:49
Operator : JC/MD
Sample : VX0518WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0518WBL01

Quant Time: May 18 02:20:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

