

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
 Data File : VX022044.D
 Acq On : 17 May 2021 20:08
 Operator : JC/MD
 Sample : M2411-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW8-GW-DUP-20210510

Quant Time: May 18 02:22:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	72845	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	158661	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	165953	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74260	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	60082	50.12	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.24%
35) Dibromofluoromethane	5.397	113	51118	47.01	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.02%
50) Toluene-d8	8.653	98	203716	49.10	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.20%
62) 4-Bromofluorobenzene	11.085	95	87092	53.18	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.36%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	7699	14.32	ug/l	99
25) 2-Butanone	4.574	43	2034	2.20	ug/l #	77
40) Benzene	6.050	78	3353	0.66	ug/l #	90
65) Chlorobenzene	10.086	112	1067	0.28	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
Data File : VX022044.D
Acq On : 17 May 2021 20:08
Operator : JC/MD
Sample : M2411-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-DUP-20210510

Quant Time: May 18 02:22:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

