

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
 Data File : VX022049.D
 Acq On : 17 May 2021 22:04
 Operator : JC/MD
 Sample : M2411-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW8-GW-858-860

Quant Time: May 18 02:24:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	70320	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	159489	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	167184	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75189	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	59034	51.01	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.02%
35) Dibromofluoromethane	5.397	113	51261	46.90	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.80%
50) Toluene-d8	8.653	98	202547	48.56	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.12%
62) 4-Bromofluorobenzene	11.085	95	87448	53.12	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.24%
Target Compounds						
16) Acetone	2.392	43	2633	5.07	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
Data File : VX022049.D
Acq On : 17 May 2021 22:04
Operator : JC/MD
Sample : M2411-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-858-860

Quant Time: May 18 02:24:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

