

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
 Data File : VX022055.D
 Acq On : 18 May 2021 11:17
 Operator : JC/MD
 Sample : VX0518WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0518WBL02

Quant Time: May 19 01:57:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	80094	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	181083	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187677	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85480	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	66687	50.59	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.18%
35) Dibromofluoromethane	5.391	113	56588	45.60	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.20%
50) Toluene-d8	8.653	98	228547	48.26	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.52%
62) 4-Bromofluorobenzene	11.085	95	99158	53.06	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.12%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
Data File : VX022055.D
Acq On : 18 May 2021 11:17
Operator : JC/MD
Sample : VX0518WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0518WBL02

Quant Time: May 19 01:57:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

