

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
 Data File : VX022070.D
 Acq On : 18 May 2021 23:16
 Operator : JC/MD
 Sample : VX0518WBL03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0518WBL03

Quant Time: May 19 09:16:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051821W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 05:14:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	71889	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	162616	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171741	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77242	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	69674	23.67	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	47.34%#
35) Dibromofluoromethane	5.397	113	59475	22.47	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	44.94%#
50) Toluene-d8	8.653	98	233192	22.96	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	45.92%#
62) 4-Bromofluorobenzene	11.085	95	99780	26.36	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	52.72%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
Data File : VX022070.D
Acq On : 18 May 2021 23:16
Operator : JC/MD
Sample : VX0518WBL03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0518WBL03

Quant Time: May 19 09:16:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051821W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 05:14:43 2021
Response via : Initial Calibration

