

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
 Data File : VX022075.D
 Acq On : 19 May 2021 01:11
 Operator : JC/MD
 Sample : M2406-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: May 19 05:32:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051821W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 05:14:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	69266	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	159181	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164455	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74383	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	58585	20.66	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	41.32%#
35) Dibromofluoromethane	5.397	113	50029	19.31	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	38.62%#
50) Toluene-d8	8.653	98	199959	20.11	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	40.22%#
62) 4-Bromofluorobenzene	11.085	95	87053	23.50	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	47.00%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
Data File : VX022075.D
Acq On : 19 May 2021 01:11
Operator : JC/MD
Sample : M2406-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: May 19 05:32:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051821W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 05:14:43 2021
Response via : Initial Calibration

