

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
 Data File : VX022076.D
 Acq On : 19 May 2021 01:35
 Operator : JC/MD
 Sample : M2406-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: May 19 05:33:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051821W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 05:14:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	69350	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	160528	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164601	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73641	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	59534	20.97	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	41.94%#
35) Dibromofluoromethane	5.397	113	50516	19.33	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	38.66%#
50) Toluene-d8	8.653	98	204495	20.39	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	40.78%#
62) 4-Bromofluorobenzene	11.085	95	86966	23.27	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	46.54%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
Data File : VX022076.D
Acq On : 19 May 2021 01:35
Operator : JC/MD
Sample : M2406-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: May 19 05:33:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051821W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 05:14:43 2021
Response via : Initial Calibration

