


cis-1,3-Dichloropropene

Response Ratio



$$\text{Response} = 6.656\text{e-}001 * \text{Amt} - 1.713\text{e-}002$$


Coef of Det (r^2) = 0.999461 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA X\Method\82X051821W.M

Calibration Table Last Updated: Wed May 19 05:14:43 2021

1,2,3-Trichlorobenzene

Response Ratio



$$\text{Response} = 1.071\text{e}+000 * \text{Amt} - 2.367\text{e}-002$$


Coef of Det (r^2) = 0.999084 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA X\Method\82X051821W.M

Calibration Table Last Updated: Wed May 19 05:14:43 2021

Naphthalene

Response Ratio



$$\text{Response} = 4.429 \times 10^0 \times \text{Amt} - 1.379 \times 10^{-1}$$


Coef of Det (r^2) = 0.999369 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA X\Method\82X051821W.M

Calibration Table Last Updated: Wed May 19 05:14:43 2021

Chloroethane

Response Ratio



$R = -9.200e-002 A^2 + 5.801e-001 A - 1.455e-002$
Coef of Det (r^2) = 0.998426 Curve Fit: Quadratic
Method Name: Z:\voasrv\HPCHEM1\MSVOA X\Method\82X051821W.M
Calibration Table Last Updated: Wed May 19 05:14:43 2021