

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
 Data File : VX022060.D
 Acq On : 18 May 2021 14:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 19 01:59:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.658	0.644	2.1	92	0.00
3 P	Chloromethane	0.791	0.721	8.8	88	0.00
4 C	Vinyl Chloride	0.913	0.766	16.1#	82	0.00
5 T	Bromomethane	0.333	0.383	-15.0	118	0.01
6 T	Chloroethane	0.475	0.501	-5.5	96	0.00
7 T	Trichlorofluoromethane	1.258	1.094	13.0	84	0.00
8 T	Diethyl Ether	0.528	0.470	11.0	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.640	0.642	-0.3	100	0.00
10 T	Methyl Iodide	0.760	0.700	7.9	83	0.00
11 T	Tert butyl alcohol	0.220	0.173	21.4	79	0.00
12 CM	1,1-Dichloroethene	0.678	0.628	7.4#	93	0.00
13 T	Acrolein	0.110	0.107	2.7	120	0.00
14 T	Allyl chloride	1.242	1.176	5.3	93	0.00
15 T	Acrylonitrile	0.413	0.384	7.0	90	0.00
16 T	Acetone	0.369	0.339	8.1	91	0.00
17 T	Carbon Disulfide	1.886	1.646	12.7	85	0.00
18 T	Methyl Acetate	1.023	1.003	2.0	97	0.00
19 T	Methyl tert-butyl Ether	2.174	2.191	-0.8	97	0.00
20 T	Methylene Chloride	0.875	0.753	13.9	97	0.00
21 T	trans-1,2-Dichloroethene	0.723	0.697	3.6	93	0.00
22 T	Diisopropyl ether	2.313	2.208	4.5	92	0.00
23 T	Vinyl Acetate	2.276	2.160	5.1	90	0.00
24 P	1,1-Dichloroethane	1.330	1.334	-0.3	98	0.00
25 T	2-Butanone	0.635	0.593	6.6	89	0.00
26 T	2,2-Dichloropropane	1.190	1.168	1.8	97	0.00
27 T	cis-1,2-Dichloroethene	0.854	0.835	2.2	95	0.00
28 T	Bromochloromethane	0.608	0.624	-2.6	98	0.00
29 T	Tetrahydrofuran	0.377	0.345	8.5	87	-0.01
30 C	Chloroform	1.356	1.337	1.4#	96	0.00
31 T	Cyclohexane	1.087	1.046	3.8	94	0.00
32 T	1,1,1-Trichloroethane	1.190	1.185	0.4	98	0.00
33 S	1,2-Dichloroethane-d4	0.823	0.827	-0.5	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.343	0.344	-0.3	98	-0.01
36 T	1,1-Dichloropropene	0.530	0.530	0.0	98	0.00
37 T	Ethyl Acetate	0.640	0.607	5.2	90	0.00
38 T	Carbon Tetrachloride	0.530	0.521	1.7	93	0.00
39 T	Methylcyclohexane	0.595	0.590	0.8	96	0.00
40 TM	Benzene	1.596	1.582	0.9	94	0.00
41 T	Methacrylonitrile	0.346	0.340	1.7	95	-0.01
42 TM	1,2-Dichloroethane	0.547	0.563	-2.9	97	0.00
43 T	Isopropyl Acetate	1.083	1.020	5.8	90	0.00
44 TM	Trichloroethene	0.388	0.396	-2.1	97	0.00
45 C	1,2-Dichloropropane	0.418	0.416	0.5#	95	0.00
46 T	Dibromomethane	0.283	0.279	1.4	96	0.00
47 T	Bromodichloromethane	0.570	0.572	-0.4	96	0.00
48 T	Methyl methacrylate	0.498	0.482	3.2	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
 Data File : VX022060.D
 Acq On : 18 May 2021 14:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 19 01:59:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.009	10.0	81	0.01
50 S	Toluene-d8	1.308	1.315	-0.5	97	0.00
51 T	4-Methyl-2-Pentanone	0.674	0.624	7.4	87	0.00
52 CM	Toluene	1.019	1.015	0.4#	94	0.00
53 T	t-1,3-Dichloropropene	0.665	0.653	1.8	94	0.00
54 T	cis-1,3-Dichloropropene	0.696	0.696	0.0	94	0.00
55 T	1,1,2-Trichloroethane	0.407	0.418	-2.7	98	0.00
56 T	Ethyl methacrylate	0.715	0.677	5.3	92	0.00
57 T	1,3-Dichloropropane	0.697	0.713	-2.3	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.350	0.354	-1.1	94	0.00
59 T	2-Hexanone	0.519	0.462	11.0	83	0.00
60 T	Dibromochloromethane	0.439	0.418	4.8	92	0.00
61 T	1,2-Dibromoethane	0.425	0.421	0.9	94	0.00
62 S	4-Bromofluorobenzene	0.516	0.507	1.7	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.322	0.366	-13.7	110	0.00
65 PM	Chlorobenzene	1.159	1.174	-1.3	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.426	0.427	-0.2	96	0.00
67 C	Ethyl Benzene	2.136	2.145	-0.4#	94	0.00
68 T	m/p-Xylenes	0.816	0.812	0.5	92	0.00
69 T	o-Xylene	0.815	0.801	1.7	95	0.00
70 T	Styrene	1.355	1.335	1.5	89	0.00
71 P	Bromoform	0.324	0.298	8.0	85	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	4.521	4.757	-5.2	93	0.00
74 T	N-amyl acetate	2.423	2.211	8.7	81	0.00
75 P	1,1,2,2-Tetrachloroethane	1.677	1.533	8.6	84	0.00
76 T	1,2,3-Trichloropropane	1.422	1.397	1.8	85	0.00
77 T	Bromobenzene	0.974	1.007	-3.4	91	0.00
78 T	n-propylbenzene	5.356	5.615	-4.8	92	0.00
79 T	2-Chlorotoluene	3.245	3.318	-2.2	91	0.00
80 T	1,3,5-Trimethylbenzene	3.885	3.948	-1.6	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.613	0.577	5.9	82	0.00
82 T	4-Chlorotoluene	3.756	3.778	-0.6	87	0.00
83 T	tert-Butylbenzene	3.766	3.787	-0.6	89	0.00
84 T	1,2,4-Trimethylbenzene	3.897	4.030	-3.4	90	0.00
85 T	sec-Butylbenzene	4.624	4.870	-5.3	91	0.00
86 T	p-Isopropyltoluene	3.941	4.155	-5.4	91	0.00
87 T	1,3-Dichlorobenzene	1.840	1.898	-3.2	91	0.00
88 T	1,4-Dichlorobenzene	1.903	1.929	-1.4	89	0.00
89 T	n-Butylbenzene	3.634	3.891	-7.1	89	0.00
90 T	Hexachloroethane	0.712	0.703	1.3	88	0.00
91 T	1,2-Dichlorobenzene	1.816	1.814	0.1	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.404	0.348	13.9	84	0.00
93 T	1,2,4-Trichlorobenzene	1.094	1.126	-2.9	92	0.00
94 T	Hexachlorobutadiene	0.427	0.433	-1.4	88	0.00
95 T	Naphthalene	4.482	4.586	-2.3	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
Data File : VX022060.D
Acq On : 18 May 2021 14:59
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 19 01:59:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.109	1.163	-4.9	92	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6