



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051821\
 Data File : VX022061.D
 Acq On : 18 May 2021 18:16
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051821W.M
 Title : SW846 8260
 Last Update : Wed May 19 05:14:43 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	5944	PASS
75	95	30	60	50.1	16116	PASS
95	95	100	100	100.0	32165	PASS
96	95	5	9	6.8	2197	PASS
173	174	0.00	2	0.9	184	PASS
174	95	50	100	65.6	21099	PASS
175	174	5	9	7.3	1549	PASS
176	174	95	101	95.2	20093	PASS
177	176	5	9	6.8	1357	PASS