

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051822\
 Data File : VX028800.D
 Acq On : 18 May 2022 12:59
 Operator : JC/MD
 Sample : N2884-04DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31767DL

Quant Time: May 19 01:33:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	261065	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	469399	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	439883	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	206193	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	180109	50.330	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.660%	
35) Dibromofluoromethane	5.385	113	156747	50.967	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.940%	
50) Toluene-d8	8.653	98	663434	57.217	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 114.440%	
62) 4-Bromofluorobenzene	11.079	95	210761	48.666	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.340%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.953	59	70851	88.126	ug/l	98
16) Acetone	2.380	43	17614	11.469	ug/l	95
25) 2-Butanone	4.562	43	8566	3.553	ug/l	96
40) Benzene	6.044	78	293523	22.110	ug/l	100
51) 4-Methyl-2-Pentanone	8.580	43	17626	3.545	ug/l	92
52) Toluene	8.720	92	336094	39.356	ug/l	100
67) Ethyl Benzene	10.195	91	262904	15.718	ug/l	98
68) m/p-Xylenes	10.299	106	126064	19.359	ug/l	99
69) o-Xylene	10.640	106	58028	9.051	ug/l	98
70) Styrene	10.659	104	107370	10.312	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	17623	1.353	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	48022	3.675	ug/l	95
95) Naphthalene	13.774	128	509333	34.762	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051822\
Data File : VX028800.D
Acq On : 18 May 2022 12:59
Operator : JC/MD
Sample : N2884-04DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31767DL

Quant Time: May 19 01:33:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

