

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051822\
 Data File : VX028801.D
 Acq On : 18 May 2022 13:22
 Operator : JC/MD
 Sample : N2884-03DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31517DL

Quant Time: May 19 01:33:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	283879	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	506180	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	480419	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	229570	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	196402	50.472	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.940%	
35) Dibromofluoromethane	5.391	113	166110	50.087	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.180%	
50) Toluene-d8	8.653	98	714937	57.179	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 114.360%	
62) 4-Bromofluorobenzene	11.085	95	230595	49.376	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 98.760%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	283813	169.942	ug/l #	1
25) 2-Butanone	4.586	43	9513	3.629	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051822\
Data File : VX028801.D
Acq On : 18 May 2022 13:22
Operator : JC/MD
Sample : N2884-03DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31517DL

Quant Time: May 19 01:33:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

