

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051822\
 Data File : VX028804.D
 Acq On : 18 May 2022 14:32
 Operator : JC/MD
 Sample : N2884-10DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31803DL

Quant Time: May 19 01:34:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	251897	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	437230	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	465707	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	180579	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	166552	48.235	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.480%
35) Dibromofluoromethane	5.385	113	134278	46.874	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.740%
50) Toluene-d8	8.653	98	511614	47.370	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.740%
62) 4-Bromofluorobenzene	11.085	95	190711	47.276	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.560%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	324090	218.697	ug/l #	1
25) 2-Butanone	4.568	43	7834	3.368	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051822\
Data File : VX028804.D
Acq On : 18 May 2022 14:32
Operator : JC/MD
Sample : N2884-10DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31803DL

Quant Time: May 19 01:34:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

