

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051823\
 Data File : VX035736.D
 Acq On : 18 May 2023 12:38
 Operator : JC/MD
 Sample : VX0518WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: May 19 05:45:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	196725	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	338602	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	295763	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	143560	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	173474	50.547	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.100%
35) Dibromofluoromethane	5.379	113	116872	51.181	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.360%
50) Toluene-d8	8.647	98	420235	49.815	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.640%
62) 4-Bromofluorobenzene	11.079	95	172375	50.749	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.500%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.167	85	49602	19.347	ug/l	100
3) Chloromethane	1.295	50	51421	16.527	ug/l	99
4) Vinyl Chloride	1.374	62	49966	18.923	ug/l	97
5) Bromomethane	1.618	94	30779	21.523	ug/l	97
6) Chloroethane	1.691	64	29751	18.573	ug/l	99
7) Trichlorofluoromethane	1.886	101	77085	19.855	ug/l	92
8) Diethyl Ether	2.130	74	28675	19.182	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.331	101	44472	19.293	ug/l	95
10) Methyl Iodide	2.453	142	41564	20.161	ug/l	# 88
11) Tert butyl alcohol	3.014	59	47058	77.719	ug/l	100
12) 1,1-Dichloroethene	2.319	96	39574	18.064	ug/l	85
13) Acrolein	2.240	56	68648	95.525	ug/l	99
14) Allyl chloride	2.666	41	81632	17.393	ug/l	93
15) Acrylonitrile	3.069	53	131419	98.567	ug/l	98
16) Acetone	2.398	43	147779	100.187	ug/l	95
17) Carbon Disulfide	2.514	76	88312	16.703	ug/l	99
18) Methyl Acetate	2.703	43	109291	20.091	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	161370	18.510	ug/l	99
20) Methylene Chloride	2.788	84	48859	18.277	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	44701	18.621	ug/l	93
22) Diisopropyl ether	3.751	45	168761	18.365	ug/l	# 84
23) Vinyl Acetate	3.721	43	604216	92.926	ug/l	98
24) 1,1-Dichloroethane	3.611	63	89903	18.293	ug/l	96
25) 2-Butanone	4.568	43	200444	98.165	ug/l	100
26) 2,2-Dichloropropane	4.471	77	73264	17.604	ug/l	97
27) cis-1,2-Dichloroethene	4.483	96	51344	17.792	ug/l	93
28) Bromochloromethane	4.898	49	43803	19.815	ug/l	95
29) Tetrahydrofuran	5.013	42	121815	94.915	ug/l	99
30) Chloroform	5.087	83	92701	18.407	ug/l	99
31) Cyclohexane	5.465	56	80617	18.592	ug/l	97
32) 1,1,1-Trichloroethane	5.379	97	79591	18.164	ug/l	99
36) 1,1-Dichloropropene	5.690	75	69105	19.172	ug/l	97
37) Ethyl Acetate	4.709	43	75202	20.003	ug/l	99
38) Carbon Tetrachloride	5.678	117	62656	18.350	ug/l	99
39) Methylcyclohexane	7.379	83	81909	19.305	ug/l	93
40) Benzene	6.038	78	188217	18.666	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051823\
 Data File : VX035736.D
 Acq On : 18 May 2023 12:38
 Operator : JC/MD
 Sample : VX0518WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: May 19 05:45:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	43590	19.528	ug/l	96
42) 1,2-Dichloroethane	6.086	62	84429	19.464	ug/l	96
43) Isopropyl Acetate	6.336	43	123099	18.777	ug/l	99
44) Trichloroethene	7.123	130	46677	18.256	ug/l	88
45) 1,2-Dichloropropane	7.428	63	49947	18.703	ug/l	95
46) Dibromomethane	7.580	93	34619	18.938	ug/l	95
47) Bromodichloromethane	7.818	83	64997	18.395	ug/l	97
48) Methyl methacrylate	7.690	41	61974	19.155	ug/l	96
49) 1,4-Dioxane	7.714	88	25444	367.827	ug/l	96
51) 4-Methyl-2-Pentanone	8.574	43	382865	101.371	ug/l	98
52) Toluene	8.714	92	118542	18.624	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	70190	18.285	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	76596	18.623	ug/l	97
55) 1,1,2-Trichloroethane	9.153	97	48285	19.314	ug/l	96
56) Ethyl methacrylate	9.116	69	75394	18.623	ug/l	94
57) 1,3-Dichloropropane	9.305	76	85449	18.855	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	200422	97.313	ug/l	98
59) 2-Hexanone	9.427	43	295106	103.752	ug/l	97
60) Dibromochloromethane	9.519	129	43421	18.221	ug/l	96
61) 1,2-Dibromoethane	9.610	107	50722	19.257	ug/l	98
64) Tetrachloroethene	9.275	164	37779	18.973	ug/l	95
65) Chlorobenzene	10.080	112	125061	19.371	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	42583	18.653	ug/l	99
67) Ethyl Benzene	10.195	91	234118	19.281	ug/l	97
68) m/p-Xylenes	10.299	106	172685	38.475	ug/l	94
69) o-Xylene	10.640	106	84656	19.008	ug/l	92
70) Styrene	10.653	104	138703	19.324	ug/l	97
71) Bromoform	10.799	173	26186	17.434	ug/l #	100
73) Isopropylbenzene	10.964	105	227013	18.766	ug/l	98
74) N-amyl acetate	10.842	43	102009	18.086	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.213	83	74392	18.957	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	79637	22.679	ug/l	83
77) Bromobenzene	11.195	156	49920	18.181	ug/l	91
78) n-propylbenzene	11.305	91	269089	18.923	ug/l	97
79) 2-Chlorotoluene	11.366	91	165352	18.572	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	196695	19.131	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	18214	16.717	ug/l #	83
82) 4-Chlorotoluene	11.451	91	190582	18.587	ug/l	97
83) tert-Butylbenzene	11.713	119	184269	18.438	ug/l	95
84) 1,2,4-Trimethylbenzene	11.750	105	193470	18.740	ug/l	96
85) sec-Butylbenzene	11.890	105	237786	18.950	ug/l	98
86) p-Isopropyltoluene	12.006	119	196069	19.006	ug/l	97
87) 1,3-Dichlorobenzene	11.969	146	97292	18.653	ug/l	98
88) 1,4-Dichlorobenzene	12.043	146	96474	18.317	ug/l	97
89) n-Butylbenzene	12.335	91	177505	18.915	ug/l	98
90) Hexachloroethane	12.536	117	26731	17.096	ug/l	93
91) 1,2-Dichlorobenzene	12.335	146	93727	18.350	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	15943	18.245	ug/l	87
93) 1,2,4-Trichlorobenzene	13.585	180	56911	18.505	ug/l	98
94) Hexachlorobutadiene	13.725	225	23893	19.573	ug/l	99
95) Naphthalene	13.774	128	201155	19.099	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	57423	18.985	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051823\
Data File : VX035736.D
Acq On : 18 May 2023 12:38
Operator : JC/MD
Sample : VX0518WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: May 19 05:45:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051823\
 Data File : VX035736.D
 Acq On : 18 May 2023 12:38
 Operator : JC/MD
 Sample : VX0518WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: May 19 05:45:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:168 Resp: 196725
Ion Ratio Lower Upper
168 100
99 71.5 45.7 68.5#




#2
Dichlorodifluoromethane
Concen: 19.347 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.001 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 85 Resp: 49602
Ion Ratio Lower Upper
85 100
87 32.9 16.4 49.1



Abundance Scan 34 (1.294 min): VX035603.D\data.ms (-29)



#3

Chloromethane

Concen: 16.527 ug/l

RT: 1.295 min Scan# 34

Delta R.T. 0.001 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38

Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion: 50 Resp: 51421

Ion Ratio Lower Upper


50 100

52 33.1 25.9 38.9


Abundance Scan 34 (1.295 min): VX035736.D\data.ms



Abundance Scan 34 (1.295 min): VX035736.D\data.ms (-1) (



Abundance Scan 47 (1.374 min): VX035603.D\data.ms (-42)



#4

Vinyl Chloride

Concen: 18.923 ug/l

RT: 1.374 min Scan# 47

Delta R.T. -0.000 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38



Tgt Ion: 62 Resp: 49966

Ion Ratio Lower Upper


62 100


64 31.4 26.4 39.6

Abundance Scan 47 (1.374 min): VX035736.D\data.ms



Abundance Scan 47 (1.374 min): VX035736.D\data.ms (-1) (









#5
Bromomethane
Concen: 21.523 ug/l
RT: 1.618 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 30779
Ion Ratio Lower Upper
94 100
96 93.0 77.0 115.6



#6
Chloroethane
Concen: 18.573 ug/l
RT: 1.691 min Scan# 99
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 64 Resp: 29751
Ion Ratio Lower Upper
64 100
66 33.1 26.2 39.4









#7
Trichlorofluoromethane
Concen: 19.855 ug/l
RT: 1.886 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 77085
Ion Ratio Lower Upper
101 100
103 58.7 52.0 78.0



#8
Diethyl Ether
Concen: 19.182 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 74 Resp: 28675
Ion Ratio Lower Upper
74 100
45 105.0 49.8 149.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.293 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 44472
Ion Ratio Lower Upper
101 100
85 46.0 35.4 53.2
151 69.8 60.2 90.4



#10
Methyl Iodide
Concen: 20.161 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion:142 Resp: 41564
Ion Ratio Lower Upper
142 100
127 51.9 34.0 51.0#
141 15.0 11.1 16.7





#11

Tert butyl alcohol

Concen: 77.719 ug/l

RT: 3.014 min Scan# 314

Delta R.T. 0.013 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 47058

Ion Ratio Lower Upper

59 100

57 10.2 8.1 12.1



#12

1,1-Dichloroethene

Concen: 18.064 ug/l

RT: 2.319 min Scan# 202

Delta R.T. -0.000 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38




Tgt Ion: 96 Resp: 39574


Ion Ratio Lower Upper

96 100

61 187.5 128.8 193.2

98 65.5 51.4 77.0









#13
Acrolein
Concen: 95.525 ug/l
RT: 2.240 min Scan# 189
Delta R.T. 0.001 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 56 Resp: 68648
Ion Ratio Lower Upper
56 100
55 71.5 56.6 85.0



#14
Allyl chloride
Concen: 17.393 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38


Tgt Ion: 41 Resp: 81632
Ion Ratio Lower Upper
41 100
39 71.7 52.4 78.6
76 29.0 26.0 39.0








#15
Acrylonitrile
Concen: 98.567 ug/l
RT: 3.069 min Scan# 31
Delta R.T. 0.001 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 53 Resp: 131419
Ion Ratio Lower Upper
53 100
52 83.5 66.2 99.2
51 38.0 28.3 42.5



#16
Acetone
Concen: 100.187 ug/l
RT: 2.398 min Scan# 215
Delta R.T. 0.006 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 43 Resp: 147779
Ion Ratio Lower Upper
43 100
58 28.2 24.7 37.1









#17
Carbon Disulfide
Concen: 16.703 ug/l
RT: 2.514 min Scan# 211
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 88312
Ion Ratio Lower Upper
76 100
78 9.4 7.2 10.8



#18
Methyl Acetate
Concen: 20.091 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 43 Resp: 109291
Ion Ratio Lower Upper
43 100
74 22.0 18.3 27.5









#19
Methyl tert-butyl Ether
Concen: 18.510 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 161370
Ion Ratio Lower Upper
73 100
57 22.7 18.5 27.7



#20
Methylene Chloride
Concen: 18.277 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 84 Resp: 48859
Ion Ratio Lower Upper
84 100
49 136.3 103.8 155.8
51 42.8 31.8 47.8
86 64.3 51.4 77.2









#21
trans-1,2-Dichloroethene
Concen: 18.621 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 96 Resp: 44701
Ion Ratio Lower Upper
96 100
61 151.3 113.0 169.6
98 61.8 52.1 78.1



#22
Diisopropyl ether
Concen: 18.365 ug/l
RT: 3.751 min Scan# 437
Delta R.T. -0.006 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 45 Resp: 168761
Ion Ratio Lower Upper
45 100
43 100.2 65.4 98.2#
87 23.0 20.6 31.0
59 10.2 9.0 13.6









#23
 Vinyl Acetate
 Concen: 92.926 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX035736.D
 Acq: 18 May 2023 12:38




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 604216
 Ion Ratio Lower Upper
 43 100
 86 9.9 8.5 12.7



#24
 1,1-Dichloroethane
 Concen: 18.293 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX035736.D
 Acq: 18 May 2023 12:38

Tgt Ion: 63 Resp: 89903
 Ion Ratio Lower Upper
 63 100
 98 5.6 3.8 11.2
 100 4.3 2.4 7.1









#25
2-Butanone
Concen: 98.165 ug/l
RT: 4.568 min Scan# 51
Delta R.T. 0.012 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 200444
Ion Ratio Lower Upper
43 100
72 24.1 19.3 28.9



#26
2,2-Dichloropropane
Concen: 17.604 ug/l
RT: 4.471 min Scan# 555
Delta R.T. -0.006 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38


Tgt Ion: 77 Resp: 73264
Ion Ratio Lower Upper
77 100
97 21.1 11.4 34.1








#27
cis-1,2-Dichloroethene
Concen: 17.792 ug/l
RT: 4.483 min Scan# 51
Delta R.T. -0.006 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 96 Resp: 51344
Ion Ratio Lower Upper
96 100
61 163.3 0.0 300.4
98 64.8 0.0 129.0



#28
Bromochloromethane
Concen: 19.815 ug/l
RT: 4.898 min Scan# 625
Delta R.T. 0.001 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 49 Resp: 43803
Ion Ratio Lower Upper
49 100
129 1.8 0.0 4.0
130 68.0 58.2 87.2





#29

Tetrahydrofuran

Concen: 94.915 ug/l

RT: 5.013 min Scan# 64

Delta R.T. 0.006 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 121815

Ion Ratio Lower Upper

42 100

72 42.2 34.5 51.7

71 39.4 32.1 48.1



#30

Chloroform

Concen: 18.407 ug/l

RT: 5.087 min Scan# 656

Delta R.T. -0.006 min

Lab File: VX035736.D




Acq: 18 May 2023 12:38


Tgt Ion: 83 Resp: 92701

Ion Ratio Lower Upper

83 100


85 62.9 50.6 76.0








#31
Cyclohexane
Concen: 18.592 ug/l
RT: 5.465 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 80617
Ion Ratio Lower Upper
56 100
69 30.4 23.7 35.5
84 82.3 68.5 102.7



#32
1,1,1-Trichloroethane
Concen: 18.164 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.006 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 97 Resp: 79591
Ion Ratio Lower Upper
97 100
99 63.8 51.1 76.7
61 47.8 36.5 54.7









#33
1,2-Dichloroethane-d4
Concen: 50.547 ug/l
RT: 5.952 min Scan# 798
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 65 Resp: 173474
Ion Ratio Lower Upper
65 100
67 50.1 0.0 104.0



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. -0.006 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38


Tgt Ion: 114 Resp: 338602
Ion Ratio Lower Upper
114 100
63 24.0 0.0 42.2
88 17.5 0.0 33.2








#35
Dibromofluoromethane
Concen: 51.181 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.006 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:113 Resp: 116872
Ion Ratio Lower Upper
113 100
111 102.1 83.7 125.5
192 16.9 15.4 23.2



#36
1,1-Dichloropropene
Concen: 19.172 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 75 Resp: 69105
Ion Ratio Lower Upper
75 100
110 32.4 17.5 52.5
77 30.0 24.3 36.5





#37

Ethyl Acetate

Concen: 20.003 ug/l

RT: 4.709 min Scan# 594

Delta R.T. -0.006 min


Lab File: VX035736.D

Acq: 18 May 2023 12:38

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 43 Resp: 75202

Ion Ratio Lower Upper

43 100

61 12.6 10.2 15.4

70 9.4 8.2 12.4



#38

Carbon Tetrachloride

Concen: 18.350 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.000 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38




Tgt Ion: 117 Resp: 62656


Ion Ratio Lower Upper

117 100

119 96.5 76.6 114.8

121 28.1 23.8 35.6





#39

Methylcyclohexane

Concen: 19.305 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. -0.000 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 81909

Ion Ratio Lower Upper

83 100

55 86.1 60.5 90.7

98 50.7 40.4 60.6



#40

Benzene

Concen: 18.666 ug/l

RT: 6.038 min Scan# 812

Delta R.T. 0.001 min

Lab File: VX035736.D




Acq: 18 May 2023 12:38


Tgt Ion: 78 Resp: 188217

Ion Ratio Lower Upper

78 100


77 23.4 19.0 28.4








#41
Methacrylonitrile
Concen: 19.528 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 41 Resp: 43590




Ion	Ratio	Lower	Upper
41	100		
39	56.0	45.0	67.6
67	63.8	55.6	83.4
52	30.5	23.2	34.8




#42
1,2-Dichloroethane
Concen: 19.464 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 62 Resp: 84429


Ion	Ratio	Lower	Upper
62	100		
98	7.5	0.0	18.0








#43
Isopropyl Acetate
Concen: 18.777 ug/l
RT: 6.336 min Scan# 861
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 123099
Ion Ratio Lower Upper
43 100
61 18.5 15.2 22.8
87 12.2 9.8 14.8



#44
Trichloroethene
Concen: 18.256 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion:130 Resp: 46677
Ion Ratio Lower Upper
130 100
95 111.2 0.0 198.4





#45

1,2-Dichloropropane

Concen: 18.703 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.001 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 49947

Ion Ratio Lower Upper

63 100

65 33.6 24.6 37.0



#46

Dibromomethane

Concen: 18.938 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.000 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38




Tgt Ion: 93 Resp: 34619


Ion Ratio Lower Upper

93 100

95 82.5 66.9 100.3

174 86.1 75.4 113.0





#47

Bromodichloromethane

Concen: 18.395 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX035736.D





Acq: 18 May 2023 12:38

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:	83	Resp:	64997
Ion	Ratio	Lower	Upper
83	100		
85	62.6	52.2	78.4
127	8.2	6.6	9.8



#48

Methyl methacrylate

Concen: 19.155 ug/l




RT: 7.690 min Scan# 1083


Delta R.T. -0.000 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38


Tgt Ion:	41	Resp:	61974
Ion	Ratio	Lower	Upper
41	100		
69	73.4	62.2	93.4
39	61.7	47.7	71.5








#49
1,4-Dioxane
Concen: 367.827 ug/l
RT: 7.714 min Scan# 1086
Delta R.T. 0.006 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 88 Resp: 25444
Ion Ratio Lower Upper
88 100
43 34.4 27.7 41.5
58 76.4 56.8 85.2



#50
Toluene-d8
Concen: 49.815 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 98 Resp: 420235
Ion Ratio Lower Upper
98 100
100 66.4 52.5 78.7









#51
4-Methyl-2-Pentanone
Concen: 101.371 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 382865
Ion Ratio Lower Upper
43 100
58 36.9 30.6 46.0



#52
Toluene
Concen: 18.624 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.006 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 92 Resp: 118542
Ion Ratio Lower Upper
92 100
91 172.6 136.8 205.2





#53

t-1,3-Dichloropropene

Concen: 18.285 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 70190

Ion Ratio Lower Upper

75 100

77 32.1 25.4 38.2



#54

cis-1,3-Dichloropropene

Concen: 18.623 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38




Tgt Ion: 75 Resp: 76596


Ion Ratio Lower Upper

75 100

77 31.5 25.0 37.4


39 59.0 44.6 66.8








#55
1,1,2-Trichloroethane
Concen: 19.314 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 97 Resp: 48285




Ion	Ratio	Lower	Upper
97	100		
83	86.3	69.7	104.5
85	57.4	43.7	65.5
99	68.3	49.7	74.5




#56
Ethyl methacrylate
Concen: 18.623 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 69 Resp: 75394

Ion	Ratio	Lower	Upper
69	100		
41	76.2	57.4	86.0
39	46.3	33.7	50.5









#57
1,3-Dichloropropane
Concen: 18.855 ug/l
RT: 9.305 min Scan# 1148
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 85449
Ion Ratio Lower Upper
76 100
78 32.8 25.8 38.8



#58
2-Chloroethyl Vinyl ether
Concen: 97.313 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 63 Resp: 200422
Ion Ratio Lower Upper
63 100
106 26.8 22.2 33.2





#59

2-Hexanone

Concen: 103.752 ug/l

RT: 9.427 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 295106

Ion Ratio Lower Upper

43 100

58 50.5 26.3 78.9



#60

Dibromochloromethane

Concen: 18.221 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX035736.D




Acq: 18 May 2023 12:38


Tgt Ion: 129 Resp: 43421

Ion Ratio Lower Upper

129 100

127 74.4 39.0 117.0









#61
1,2-Dibromoethane
Concen: 19.257 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 50722
Ion Ratio Lower Upper
107 100
109 92.3 75.3 112.9



#62
4-Bromofluorobenzene
Concen: 50.749 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38


Tgt Ion: 95 Resp: 172375
Ion Ratio Lower Upper
95 100
174 65.9 0.0 147.4
176 62.3 0.0 140.6








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:117 Resp: 295763
Ion Ratio Lower Upper
117 100
82 61.8 45.0 67.4
119 31.9 26.0 39.0




#64
Tetrachloroethene
Concen: 18.973 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38



Tgt Ion:164 Resp: 37779
Ion Ratio Lower Upper
164 100
166 128.4 102.7 154.1
129 101.1 75.6 113.4
131 103.0 73.7 110.5









#65
Chlorobenzene
Concen: 19.371 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.001 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 125061
Ion Ratio Lower Upper
112 100
114 33.6 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 18.653 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion:131 Resp: 42583
Ion Ratio Lower Upper
131 100
133 96.2 47.1 141.4
119 65.6 32.6 98.0





#67

Ethyl Benzene

Concen: 19.281 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 234118

Ion Ratio Lower Upper

91 100

106 30.5 25.6 38.4



#68

m/p-Xylenes

Concen: 38.475 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX035736.D




Acq: 18 May 2023 12:38


Tgt Ion:106 Resp: 172685

Ion Ratio Lower Upper

106 100


91 214.3 164.0 246.0








#69
o-Xylene
Concen: 19.008 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:106 Resp: 84656
Ion Ratio Lower Upper
106 100
91 229.9 108.5 325.3



#70
Styrene
Concen: 19.324 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38


Tgt Ion:104 Resp: 138703
Ion Ratio Lower Upper
104 100
78 58.2 43.4 65.2
103 54.6 44.2 66.2








#71
Bromoform
Concen: 17.434 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:173 Resp: 26186
Ion Ratio Lower Upper
173 100
175 48.0 24.1 72.4
254 0.0 0.1 0.1#




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38



Tgt Ion:152 Resp: 143560
Ion Ratio Lower Upper
152 100
115 76.2 45.0 135.0
150 164.8 0.0 352.4









#73
Isopropylbenzene
Concen: 18.766 ug/l
RT: 10.964 min Scan# 1620
Delta R.T. 0.001 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 227013
Ion Ratio Lower Upper
105 100
120 25.4 13.3 39.9



#74
N-ethyl acetate
Concen: 18.086 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.001 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 43 Resp: 102009
Ion Ratio Lower Upper
43 100
70 39.0 32.1 48.1
55 22.5 18.3 27.5
61 22.4 18.8 28.2









#75
1,1,2,2-Tetrachloroethane
Concen: 18.957 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 74392
Ion Ratio Lower Upper
83 100
131 9.8 4.8 14.3
85 62.7 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 22.679 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38


Tgt Ion: 75 Resp: 79637
Ion Ratio Lower Upper
75 100
77 32.3 21.6 64.8








#77
Bromobenzene
Concen: 18.181 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:156 Resp: 49920
Ion Ratio Lower Upper
156 100
77 192.3 86.7 259.9
158 95.6 48.0 144.0




#78
n-propylbenzene
Concen: 18.923 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38



Tgt Ion: 91 Resp: 269089
Ion Ratio Lower Upper
91 100
120 21.7 11.5 34.5









#79
2-Chlorotoluene
Concen: 18.572 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 165352
Ion Ratio Lower Upper
91 100
126 30.9 16.3 48.9



#80
1,3,5-Trimethylbenzene
Concen: 19.131 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38


Tgt Ion:105 Resp: 196695
Ion Ratio Lower Upper
105 100
120 46.2 23.7 71.1








#81
trans-1,4-Dichloro-2-butene
Concen: 16.717 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 75 Resp: 18214
Ion Ratio Lower Upper
75 100
53 128.0 82.6 124.0#
89 44.8 36.9 55.3



#82
4-Chlorotoluene
Concen: 18.587 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion: 91 Resp: 190582
Ion Ratio Lower Upper
91 100
126 26.8 14.2 42.6









#83
tert-Butylbenzene
Concen: 18.438 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:119 Resp: 184269
Ion Ratio Lower Upper
119 100
91 66.3 30.6 92.0
134 23.4 11.8 35.4



#84
1,2,4-Trimethylbenzene
Concen: 18.740 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion:105 Resp: 193470
Ion Ratio Lower Upper
105 100
120 42.3 22.4 67.0









#85
 sec-Butylbenzene
 Concen: 18.950 ug/l
 RT: 11.890 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX035736.D
 Acq: 18 May 2023 12:38




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:105 Resp: 237786
 Ion Ratio Lower Upper
 105 100
 134 19.3 10.2 30.4



#86
 p-Isopropyltoluene
 Concen: 19.006 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX035736.D
 Acq: 18 May 2023 12:38

Tgt Ion:119 Resp: 196069
 Ion Ratio Lower Upper
 119 100
 134 25.0 13.2 39.5
 91 27.9 12.9 38.7





#87

1,3-Dichlorobenzene

Concen: 18.653 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 97292

Ion Ratio Lower Upper

146 100

111 44.0 20.9 62.8

148 62.5 31.6 95.0



#88

1,4-Dichlorobenzene

Concen: 18.317 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. 0.001 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38




Tgt Ion:146 Resp: 96474


Ion Ratio Lower Upper

146 100

111 44.7 20.6 61.9

148 65.1 31.8 95.4





#89

n-Butylbenzene

Concen: 18.915 ug/l

RT: 12.335 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX035736.D

Acq: 18 May 2023 12:38

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 177505

Ion Ratio Lower Upper

91 100

92 54.4 27.7 83.0

134 24.2 13.0 38.9



#90

Hexachloroethane

Concen: 17.096 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX035736.D




Acq: 18 May 2023 12:38


Tgt Ion:117 Resp: 26731

Ion Ratio Lower Upper

117 100


201 62.9 34.4 103.1








#91
1,2-Dichlorobenzene
Concen: 18.350 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:146 Resp: 93727
Ion Ratio Lower Upper
146 100
111 47.1 22.0 66.0
148 64.1 32.0 96.0




#92
1,2-Dibromo-3-Chloropropane
Concen: 18.245 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Tgt Ion: 75 Resp: 15943
Ion Ratio Lower Upper
75 100
155 66.4 38.4 115.1
157 85.6 49.9 149.7








#93
1,2,4-Trichlorobenzene
Concen: 18.505 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:180 Resp: 56911
Ion Ratio Lower Upper
180 100
182 94.9 47.0 141.0
145 34.8 15.9 47.7




#94
Hexachlorobutadiene
Concen: 19.573 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38




Tgt Ion:225 Resp: 23893
Ion Ratio Lower Upper
225 100
223 63.0 32.1 96.5
227 64.4 31.8 95.4








#95
Naphthalene
Concen: 19.099 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:128 Resp: 201155
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.4
129 11.0 8.6 13.0



#96
1,2,3-Trichlorobenzene
Concen: 18.985 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX035736.D
Acq: 18 May 2023 12:38

Tgt Ion:180 Resp: 57423
Ion Ratio Lower Upper
180 100
182 94.4 48.0 144.2
145 36.2 17.7 53.1

