

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX051823\  
 Data File : VX035733.D  
 Acq On : 18 May 2023 10:57  
 Operator : JC/MD  
 Sample : VX0518MBL01  
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_X/MEOH  
 ALS Vial : 3 Sample Multiplier: 1

**Instrument :**  
 MSVOA\_X  
**ClientSampleId :**  
 VX0518MBL01

Quant Time: May 19 05:45:15 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051123W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 11 13:57:54 2023  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.550	168	192689	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	354621	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	328106	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	150990	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	161039	47.906	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.820%
35) Dibromofluoromethane	5.379	113	115582	48.329	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.660%
50) Toluene-d8	8.647	98	429438	48.606	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.220%
62) 4-Bromofluorobenzene	11.079	95	176948	49.742	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX051823\  
 Data File : VX035733.D  
 Acq On : 18 May 2023 10:57  
 Operator : JC/MD  
 Sample : VX0518MBL01  
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_X/MEOH  
 ALS Vial : 3 Sample Multiplier: 1

**Instrument :**  
 MSVOA\_X  
**ClientSampleId :**  
 VX0518MBL01

Quant Time: May 19 05:45:15 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051123W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 11 13:57:54 2023  
 Response via : Initial Calibration

