

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX051823\  
 Data File : VX035749.D  
 Acq On : 18 May 2023 17:55  
 Operator : JC/MD  
 Sample : 02816-02ME  
 Misc : 4.18g/10mL/100uL/5.00mL/MSVOA\_X/MEOH  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 CB-22-K2-SO-8-051623ME

Quant Time: May 19 05:47:22 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051123W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 11 13:57:54 2023  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	5.550	168	188237	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	348566	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	313448	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136443	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	156898	47.778	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.560%
35) Dibromofluoromethane	5.385	113	103352	43.966	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.940%
50) Toluene-d8	8.647	98	419501	48.307	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.620%
62) 4-Bromofluorobenzene	11.085	95	167830	47.998	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.000%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	4755	2.156	ug/l	95
27) cis-1,2-Dichloroethene	4.483	96	17575	6.365	ug/l	95
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX051823\  
 Data File : VX035749.D  
 Acq On : 18 May 2023 17:55  
 Operator : JC/MD  
 Sample : 02816-02ME  
 Misc : 4.18g/10mL/100uL/5.00mL/MSVOA\_X/MEOH  
 ALS Vial : 19 Sample Multiplier: 1

**Instrument :**  
 MSVOA\_X  
**ClientSampleId :**  
 CB-22-K2-SO-8-051623ME

Quant Time: May 19 05:47:22 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051123W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 11 13:57:54 2023  
 Response via : Initial Calibration

