



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051823\
Data File : VX035731.D
Acq On : 18 May 2023 08:31
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Title : SW846 8260
Last Update : Thu May 11 13:57:54 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	6172	PASS
75	95	30	60	55.1	16615	PASS
95	95	100	100	100.0	30163	PASS
96	95	5	9	6.8	2059	PASS
173	174	0.00	2	0.9	177	PASS
174	95	50	100	66.0	19893	PASS
175	174	5	9	7.5	1487	PASS
176	174	95	101	97.4	19380	PASS
177	176	5	9	6.8	1309	PASS