

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051920\
 Data File : VX016326.D
 Acq On : 19 May 2020 12:03
 Operator : JC/SP
 Sample : L2672-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP021MW352-200514

Quant Time: May 20 04:20:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	253773	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	422493	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	401014	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	201800	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	147419	44.15	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	88.30%	
35) Dibromofluoromethane	5.47	113	121009	45.21	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	90.42%	
50) Toluene-d8	8.70	98	517208	50.15	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	100.30%	
62) 4-Bromofluorobenzene	11.12	95	202887	51.72	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	103.44%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	1172	0.352	ug/l	98
12) 1,1-Dichloroethene	2.36	96	11478	4.351	ug/l	95
16) Acetone	2.42	43	4672	3.136	ug/l	92
21) trans-1,2-Dichloroethene	3.15	96	12575	4.278	ug/l	93
27) cis-1,2-Dichloroethene	4.57	96	263097	81.864	ug/l	85
30) Chloroform	5.18	83	1858	0.366	ug/l	89
32) 1,1,1-Trichloroethane	5.47	97	28179	6.100	ug/l	99
42) 1,2-Dichloroethane	6.16	62	6921	1.579	ug/l	100
44) Trichloroethene	7.19	130	47197	10.918	ug/l	93
64) Tetrachloroethene	9.32	164	2689637	546.345	ug/l	96
94) Hexachlorobutadiene	13.76	225	1625	1.294	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051920\
Data File : VX016326.D
Acq On : 19 May 2020 12:03
Operator : JC/SP
Sample : L2672-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP021MW352-200514

Quant Time: May 20 04:20:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

