

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051920\
 Data File : VX016327.D
 Acq On : 19 May 2020 12:26
 Operator : JC/SP
 Sample : L2672-13DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP021MW352-200514DL

Quant Time: May 20 04:25:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	226608	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	378247	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	378939	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	202137	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	135405	45.41	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	90.82%	
35) Dibromofluoromethane	5.46	113	114118	47.62	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.24%	
50) Toluene-d8	8.70	98	461559	49.99	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.98%	
62) 4-Bromofluorobenzene	11.12	95	198491	56.52	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	113.04%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
12) 1,1-Dichloroethene	2.35	96	1563	0.664	ug/l	89
21) trans-1,2-Dichloroethene	3.14	96	1670	0.636	ug/l	88
27) cis-1,2-Dichloroethene	4.57	96	25739	8.969	ug/l	85
32) 1,1,1-Trichloroethane	5.46	97	2848	0.690	ug/l #	50
44) Trichloroethene	7.18	130	4920	1.271	ug/l	89
64) Tetrachloroethene	9.32	164	272854	58.654	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051920\
Data File : VX016327.D
Acq On : 19 May 2020 12:26
Operator : JC/SP
Sample : L2672-13DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP021MW352-200514DL

Quant Time: May 20 04:25:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

