

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051920\
 Data File : VX016335.D
 Acq On : 19 May 2020 15:28
 Operator : JC/SP
 Sample : L2672-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP021MW174-200514

Quant Time: May 20 06:39:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	232396	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	380717	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	377464	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	194179	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	137099	44.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.66%	
35) Dibromofluoromethane	5.47	113	110206	45.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.38%	
50) Toluene-d8	8.70	98	472802	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
62) 4-Bromofluorobenzene	11.12	95	190867	54.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.00%	
Target Compounds						
16) Acetone	2.42	43	2217	1.625	ug/l #	87
27) cis-1,2-Dichloroethene	4.56	96	1244	0.423	ug/l	63
65) Chlorobenzene	10.12	112	4426	0.560	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051920\
Data File : VX016335.D
Acq On : 19 May 2020 15:28
Operator : JC/SP
Sample : L2672-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP021MW174-200514

Quant Time: May 20 06:39:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

