

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051920\
 Data File : VX016340.D
 Acq On : 19 May 2020 17:18
 Operator : JC/SP
 Sample : L2663-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1378

Quant Time: May 20 07:23:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	144460	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	165630	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	90801	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	23676	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	25526	13.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	26.86%	
35) Dibromofluoromethane	5.47	113	29364	27.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	55.96%	
50) Toluene-d8	8.70	98	180815	44.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.44%	
62) 4-Bromofluorobenzene	11.12	95	31990	20.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	41.60%	
Target Compounds						
16) Acetone	2.42	43	26905	31.721	ug/l	97
25) 2-Butanone	4.65	43	5790	4.236	ug/l #	89
30) Chloroform	5.19	83	1103	0.381	ug/l	99
51) 4-Methyl-2-Pentanone	8.62	43	6017	3.261	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051920\
Data File : VX016340.D
Acq On : 19 May 2020 17:18
Operator : JC/SP
Sample : L2663-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1378

Quant Time: May 20 07:23:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

