


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX051920\
Data File : VX016321.D
Acq On : 19 May 2020 09:29
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Title : SW846 8260
Last Update : Wed May 06 06:43:32 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	9585	PASS
75	95	30	60	47.3	28115	PASS
95	95	100	100	100.0	59496	PASS
96	95	5	9	6.4	3811	PASS
173	174	0.00	2	0.7	351	PASS
174	95	50	100	80.5	47888	PASS
175	174	5	9	7.1	3421	PASS
176	174	95	101	97.4	46621	PASS
177	176	5	9	6.2	2898	PASS