

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051921\
 Data File : VX022087.D
 Acq On : 19 May 2021 14:41
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX051921

Quant Time: May 20 04:22:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.608	0.623	-2.5	99	0.00
3 P	Chloromethane	0.656	0.647	1.4	97	0.00
4 C	Vinyl Chloride	0.694	0.711	-2.4#	100	0.00
5 T	Bromomethane	0.404	0.377	6.7	99	0.00
6 T	Chloroethane	0.413	0.439	-6.3	102	0.00
7 T	Trichlorofluoromethane	0.998	1.021	-2.3	101	0.00
8 T	Diethyl Ether	0.429	0.428	0.2	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.592	0.592	0.0	100	0.00
10 T	Methyl Iodide	0.637	0.584	8.3	88	0.00
11 T	Tert butyl alcohol	0.176	0.168	4.5	98	-0.01
12 CM	1,1-Dichloroethene	0.573	0.583	-1.7#	101	0.00
13 T	Acrolein	0.117	0.095	18.8	88	0.00
14 T	Allyl chloride	1.053	1.061	-0.8	100	0.00
15 T	Acrylonitrile	0.347	0.349	-0.6	99	0.00
16 T	Acetone	0.336	0.322	4.2	100	0.00
17 T	Carbon Disulfide	1.501	1.451	3.3	101	0.00
18 T	Methyl Acetate	0.893	0.855	4.3	99	0.00
19 T	Methyl tert-butyl Ether	1.963	1.984	-1.1	100	0.00
20 T	Methylene Chloride	0.746	0.683	8.4	99	0.00
21 T	trans-1,2-Dichloroethene	0.623	0.635	-1.9	99	0.00
22 T	Diisopropyl ether	2.027	2.015	0.6	99	0.00
23 T	Vinyl Acetate	1.952	1.963	-0.6	99	0.00
24 P	1,1-Dichloroethane	1.190	1.201	-0.9	100	0.00
25 T	2-Butanone	0.543	0.545	-0.4	99	0.00
26 T	2,2-Dichloropropane	1.056	1.072	-1.5	102	0.00
27 T	cis-1,2-Dichloroethene	0.758	0.773	-2.0	99	0.00
28 T	Bromochloromethane	0.577	0.559	3.1	100	-0.01
29 T	Tetrahydrofuran	0.323	0.312	3.4	97	0.00
30 C	Chloroform	1.229	1.223	0.5#	99	0.00
31 T	Cyclohexane	0.969	0.937	3.3	98	-0.01
32 T	1,1,1-Trichloroethane	1.050	1.068	-1.7	101	0.00
33 S	1,2-Dichloroethane-d4	0.771	0.763	1.0	109	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.317	0.312	1.6	106	0.00
36 T	1,1-Dichloropropene	0.462	0.475	-2.8	100	0.00
37 T	Ethyl Acetate	0.567	0.553	2.5	99	0.00
38 T	Carbon Tetrachloride	0.452	0.476	-5.3	98	0.00
39 T	Methylcyclohexane	0.524	0.528	-0.8	101	0.00
40 TM	Benzene	1.423	1.437	-1.0	99	0.00
41 T	Methacrylonitrile	0.303	0.296	2.3	97	0.00
42 TM	1,2-Dichloroethane	0.498	0.512	-2.8	101	0.00
43 T	Isopropyl Acetate	0.927	0.926	0.1	99	0.00
44 TM	Trichloroethene	0.354	0.360	-1.7	100	0.00
45 C	1,2-Dichloropropane	0.363	0.377	-3.9#	98	0.00
46 T	Dibromomethane	0.246	0.254	-3.3	97	0.00
47 T	Bromodichloromethane	0.493	0.508	-3.0	97	0.00
48 T	Methyl methacrylate	0.427	0.432	-1.2	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051921\
 Data File : VX022087.D
 Acq On : 19 May 2021 14:41
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX051921

Quant Time: May 20 04:22:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	97	-0.01
50 S	Toluene-d8	1.221	1.217	0.3	107	0.00
51 T	4-Methyl-2-Pentanone	0.555	0.569	-2.5	98	0.00
52 CM	Toluene	0.901	0.906	-0.6#	98	0.00
53 T	t-1,3-Dichloropropene	0.569	0.589	-3.5	100	0.00
54 T	cis-1,3-Dichloropropene	0.611	0.629	-2.9	99	0.00
55 T	1,1,2-Trichloroethane	0.366	0.371	-1.4	98	0.00
56 T	Ethyl methacrylate	0.591	0.613	-3.7	98	0.00
57 T	1,3-Dichloropropane	0.624	0.635	-1.8	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.320	0.315	1.6	98	0.00
59 T	2-Hexanone	0.418	0.430	-2.9	98	0.00
60 T	Dibromochloromethane	0.356	0.381	-7.0	100	0.00
61 T	1,2-Dibromoethane	0.364	0.384	-5.5	100	0.00
62 S	4-Bromofluorobenzene	0.451	0.459	-1.8	109	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.328	0.335	-2.1	98	0.00
65 PM	Chlorobenzene	1.041	1.070	-2.8	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.362	0.387	-6.9	100	0.00
67 C	Ethyl Benzene	1.915	1.980	-3.4#	100	0.00
68 T	m/p-Xylenes	0.724	0.738	-1.9	100	0.00
69 T	o-Xylene	0.704	0.731	-3.8	101	0.00
70 T	Styrene	1.153	1.214	-5.3	99	0.00
71 P	Bromoform	0.250	0.276	-10.4	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	4.232	4.196	0.9	100	0.00
74 T	N-amyl acetate	2.009	1.996	0.6	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.403	1.382	1.5	99	0.00
76 T	1,2,3-Trichloropropane	1.234	1.270	-2.9	100	0.00
77 T	Bromobenzene	0.897	0.906	-1.0	100	0.00
78 T	n-propylbenzene	4.957	5.006	-1.0	101	0.00
79 T	2-Chlorotoluene	3.006	2.884	4.1	98	0.00
80 T	1,3,5-Trimethylbenzene	3.507	3.518	-0.3	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.521	0.513	1.5	96	0.00
82 T	4-Chlorotoluene	3.464	3.393	2.0	101	0.00
83 T	tert-Butylbenzene	3.364	3.370	-0.2	100	0.00
84 T	1,2,4-Trimethylbenzene	3.530	3.582	-1.5	100	0.00
85 T	sec-Butylbenzene	4.275	4.358	-1.9	100	0.00
86 T	p-Isopropyltoluene	3.610	3.711	-2.8	102	0.00
87 T	1,3-Dichlorobenzene	1.730	1.730	0.0	100	0.00
88 T	1,4-Dichlorobenzene	1.788	1.731	3.2	100	0.00
89 T	n-Butylbenzene	3.340	3.497	-4.7	102	0.00
90 T	Hexachloroethane	0.597	0.608	-1.8	97	0.00
91 T	1,2-Dichlorobenzene	1.662	1.657	0.3	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.300	0.316	-5.3	104	0.00
93 T	1,2,4-Trichlorobenzene	1.003	1.056	-5.3	106	0.00
94 T	Hexachlorobutadiene	0.391	0.400	-2.3	102	0.00
95 T	Naphthalene	3.910	4.112	-5.2	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051921\
Data File : VX022087.D
Acq On : 19 May 2021 14:41
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX051921

Quant Time: May 20 04:22:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.011	1.047	-3.6	104	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6