

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051921\
 Data File : VX022101.D
 Acq On : 19 May 2021 20:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 20 04:24:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.608	0.604	0.7	90	0.00
3 P	Chloromethane	0.656	0.662	-0.9	92	0.00
4 C	Vinyl Chloride	0.694	0.695	-0.1#	91	0.00
5 T	Bromomethane	0.404	0.383	5.2	93	0.00
6 T	Chloroethane	0.413	0.443	-7.3	95	0.00
7 T	Trichlorofluoromethane	0.998	1.014	-1.6	93	0.00
8 T	Diethyl Ether	0.429	0.440	-2.6	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.592	0.581	1.9	91	0.00
10 T	Methyl Iodide	0.637	0.647	-1.6	91	0.00
11 T	Tert butyl alcohol	0.176	0.189	-7.4	103	0.00
12 CM	1,1-Dichloroethene	0.573	0.584	-1.9#	94	0.00
13 T	Acrolein	0.117	0.101	13.7	87	0.00
14 T	Allyl chloride	1.053	1.072	-1.8	94	0.00
15 T	Acrylonitrile	0.347	0.375	-8.1	99	0.00
16 T	Acetone	0.336	0.335	0.3	97	0.00
17 T	Carbon Disulfide	1.501	1.410	6.1	91	0.00
18 T	Methyl Acetate	0.893	0.933	-4.5	100	0.00
19 T	Methyl tert-butyl Ether	1.963	2.053	-4.6	96	0.00
20 T	Methylene Chloride	0.746	0.707	5.2	95	0.00
21 T	trans-1,2-Dichloroethene	0.623	0.656	-5.3	95	0.00
22 T	Diisopropyl ether	2.027	2.082	-2.7	95	0.00
23 T	Vinyl Acetate	1.952	2.053	-5.2	96	0.00
24 P	1,1-Dichloroethane	1.190	1.244	-4.5	96	0.00
25 T	2-Butanone	0.543	0.584	-7.6	99	0.00
26 T	2,2-Dichloropropane	1.056	0.939	11.1	83	0.00
27 T	cis-1,2-Dichloroethene	0.758	0.782	-3.2	93	0.00
28 T	Bromochloromethane	0.577	0.577	0.0	96	0.00
29 T	Tetrahydrofuran	0.323	0.338	-4.6	98	0.00
30 C	Chloroform	1.229	1.248	-1.5#	94	0.00
31 T	Cyclohexane	0.969	0.929	4.1	90	0.00
32 T	1,1,1-Trichloroethane	1.050	1.078	-2.7	94	0.00
33 S	1,2-Dichloroethane-d4	0.771	0.757	1.8	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.317	0.301	5.0	98	0.00
36 T	1,1-Dichloropropene	0.462	0.462	0.0	94	0.00
37 T	Ethyl Acetate	0.567	0.570	-0.5	98	0.00
38 T	Carbon Tetrachloride	0.452	0.471	-4.2	94	0.00
39 T	Methylcyclohexane	0.524	0.495	5.5	91	0.00
40 TM	Benzene	1.423	1.426	-0.2	95	0.00
41 T	Methacrylonitrile	0.303	0.303	0.0	96	0.00
42 TM	1,2-Dichloroethane	0.498	0.507	-1.8	96	0.00
43 T	Isopropyl Acetate	0.927	0.941	-1.5	97	0.00
44 TM	Trichloroethene	0.354	0.356	-0.6	95	0.00
45 C	1,2-Dichloropropane	0.363	0.374	-3.0#	93	0.00
46 T	Dibromomethane	0.246	0.253	-2.8	94	0.00
47 T	Bromodichloromethane	0.493	0.511	-3.7	94	0.00
48 T	Methyl methacrylate	0.427	0.447	-4.7	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051921\
 Data File : VX022101.D
 Acq On : 19 May 2021 20:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 20 04:24:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	101	0.00
50 S	Toluene-d8	1.221	1.134	7.1	96	0.00
51 T	4-Methyl-2-Pentanone	0.555	0.593	-6.8	98	0.00
52 CM	Toluene	0.901	0.893	0.9#	93	0.00
53 T	t-1,3-Dichloropropene	0.569	0.573	-0.7	93	0.00
54 T	cis-1,3-Dichloropropene	0.611	0.621	-1.6	95	0.00
55 T	1,1,2-Trichloroethane	0.366	0.381	-4.1	97	0.00
56 T	Ethyl methacrylate	0.591	0.620	-4.9	95	0.00
57 T	1,3-Dichloropropane	0.624	0.646	-3.5	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.320	0.321	-0.3	96	0.00
59 T	2-Hexanone	0.418	0.448	-7.2	98	0.00
60 T	Dibromochloromethane	0.356	0.378	-6.2	95	0.00
61 T	1,2-Dibromoethane	0.364	0.390	-7.1	97	0.00
62 S	4-Bromofluorobenzene	0.451	0.431	4.4	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.328	0.328	0.0	94	0.00
65 PM	Chlorobenzene	1.041	1.037	0.4	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.362	0.379	-4.7	95	0.00
67 C	Ethyl Benzene	1.915	1.918	-0.2#	95	0.00
68 T	m/p-Xylenes	0.724	0.718	0.8	95	0.00
69 T	o-Xylene	0.704	0.706	-0.3	95	0.00
70 T	Styrene	1.153	1.192	-3.4	95	0.00
71 P	Bromoform	0.250	0.268	-7.2	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	4.232	4.155	1.8	94	0.00
74 T	N-amyl acetate	2.009	2.065	-2.8	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.403	1.432	-2.1	98	0.00
76 T	1,2,3-Trichloropropane	1.234	1.277	-3.5	95	0.00
77 T	Bromobenzene	0.897	0.887	1.1	93	0.00
78 T	n-propylbenzene	4.957	4.880	1.6	93	0.00
79 T	2-Chlorotoluene	3.006	2.889	3.9	94	0.00
80 T	1,3,5-Trimethylbenzene	3.507	3.532	-0.7	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.521	0.512	1.7	91	0.00
82 T	4-Chlorotoluene	3.464	3.333	3.8	94	0.00
83 T	tert-Butylbenzene	3.364	3.314	1.5	93	0.00
84 T	1,2,4-Trimethylbenzene	3.530	3.550	-0.6	94	0.00
85 T	sec-Butylbenzene	4.275	4.245	0.7	93	0.00
86 T	p-Isopropyltoluene	3.610	3.552	1.6	93	0.00
87 T	1,3-Dichlorobenzene	1.730	1.707	1.3	94	0.00
88 T	1,4-Dichlorobenzene	1.788	1.707	4.5	94	0.00
89 T	n-Butylbenzene	3.340	3.294	1.4	91	0.00
90 T	Hexachloroethane	0.597	0.596	0.2	90	0.00
91 T	1,2-Dichlorobenzene	1.662	1.666	-0.2	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.300	0.322	-7.3	101	0.00
93 T	1,2,4-Trichlorobenzene	1.003	0.984	1.9	94	0.00
94 T	Hexachlorobutadiene	0.391	0.367	6.1	89	0.00
95 T	Naphthalene	3.910	4.120	-5.4	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051921\
Data File : VX022101.D
Acq On : 19 May 2021 20:28
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 20 04:24:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.011	1.009	0.2	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6