

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051922\
 Data File : VX028815.D
 Acq On : 19 May 2022 11:09
 Operator : JC/MD
 Sample : VX0519WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0519WBL01

Quant Time: May 19 17:52:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	241299	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	438397	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	411448	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	185985	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	169561	51.263	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.520%	
35) Dibromofluoromethane	5.385	113	145952	50.813	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.620%	
50) Toluene-d8	8.647	98	523673	48.358	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.720%	
62) 4-Bromofluorobenzene	11.079	95	190823	47.178	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 94.360%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051922\
Data File : VX028815.D
Acq On : 19 May 2022 11:09
Operator : JC/MD
Sample : VX0519WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0519WBL01

Quant Time: May 19 17:52:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

