

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051922\
 Data File : VX028818.D
 Acq On : 19 May 2022 12:26
 Operator : JC/MD
 Sample : N2941-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 12BR-20220518

Quant Time: May 20 01:45:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	264958	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	480290	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	549156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	254008	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	184787	50.878	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.760%	
35) Dibromofluoromethane	5.391	113	158251	50.290	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.580%	
50) Toluene-d8	8.653	98	631874	53.260	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.520%	
62) 4-Bromofluorobenzene	11.079	95	274661	61.982	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 123.960%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	4768	3.059	ug/l	95
27) cis-1,2-Dichloroethene	4.489	96	2335	0.647	ug/l	94
44) Trichloroethene	7.129	130	1058	0.287	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051922\
Data File : VX028818.D
Acq On : 19 May 2022 12:26
Operator : JC/MD
Sample : N2941-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
12BR-20220518

Quant Time: May 20 01:45:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

