

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051922\
 Data File : VX028827.D
 Acq On : 19 May 2022 15:56
 Operator : JC/MD
 Sample : N2941-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 14BR-20220518

Quant Time: May 20 01:48:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	256656	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	458940	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	439151	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	201055	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	179742	51.090	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.180%	
35) Dibromofluoromethane	5.385	113	155550	51.731	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.460%	
50) Toluene-d8	8.653	98	557316	49.161	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.320%	
62) 4-Bromofluorobenzene	11.079	95	205345	48.496	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.000%	
Target Compounds						
16) Acetone	2.380	43	3386	2.243	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051922\
Data File : VX028827.D
Acq On : 19 May 2022 15:56
Operator : JC/MD
Sample : N2941-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
14BR-20220518

Quant Time: May 20 01:48:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

