

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051922\
 Data File : VX028829.D
 Acq On : 19 May 2022 16:42
 Operator : JC/MD
 Sample : N2879-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP1-0516

Quant Time: May 19 17:52:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	270189	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	485888	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	475652	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	217751	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	186176	50.268	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 100.540%		
35) Dibromofluoromethane	5.391	113	161544	50.745	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 101.480%		
50) Toluene-d8	8.653	98	587358	48.937	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 97.880%		
62) 4-Bromofluorobenzene	11.079	95	226395	50.502	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 101.000%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	26074	16.404	ug/l	95
20) Methylene Chloride	2.794	84	17537	5.259	ug/l	97
25) 2-Butanone	4.574	43	5173	2.073	ug/l	90
43) Isopropyl Acetate	6.348	43	19094	2.444	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051922\
Data File : VX028829.D
Acq On : 19 May 2022 16:42
Operator : JC/MD
Sample : N2879-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP1-0516

Quant Time: May 19 17:52:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

