



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051922\
 Data File : VX028812.D
 Acq On : 19 May 2022 09:42
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Title : SW846 8260
 Last Update : Thu May 12 14:58:53 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	11817	PASS
75	95	30	60	53.6	33037	PASS
95	95	100	100	100.0	61637	PASS
96	95	5	9	6.6	4063	PASS
173	174	0.00	2	1.3	597	PASS
174	95	50	100	73.2	45147	PASS
175	174	5	9	7.8	3515	PASS
176	174	95	101	97.5	44019	PASS
177	176	5	9	6.4	2834	PASS