

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051923\
 Data File : VX035754.D
 Acq On : 19 May 2023 12:47
 Operator : JC/MD
 Sample : VX0519MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0519MBL01

Quant Time: May 22 05:14:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	195480	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	370355	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	311015	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	130885	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	178964	52.479	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.960%	
35) Dibromofluoromethane	5.385	113	113557	45.465	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 90.940%	
50) Toluene-d8	8.647	98	411762	44.626	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 89.260%#	
62) 4-Bromofluorobenzene	11.079	95	154337	41.542	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 83.080%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051923\
Data File : VX035754.D
Acq On : 19 May 2023 12:47
Operator : JC/MD
Sample : VX0519MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0519MBL01

Quant Time: May 22 05:14:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

