

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051923\
 Data File : VX035761.D
 Acq On : 19 May 2023 15:31
 Operator : JC/MD
 Sample : 02860-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-PRO-10-131-05152023

Quant Time: May 22 05:17:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	133081	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	241147	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204028	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	123232	53.080	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.160%
35) Dibromofluoromethane	5.385	113	75623	46.500	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.000%
50) Toluene-d8	8.647	98	281904	46.922	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.840%
62) 4-Bromofluorobenzene	11.079	95	101915	42.130	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.260%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.319	96	513	0.346	ug/l	# 70
16) Acetone	2.404	43	6202	6.215	ug/l	96
44) Trichloroethene	7.123	130	324063	177.968	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051923\
Data File : VX035761.D
Acq On : 19 May 2023 15:31
Operator : JC/MD
Sample : 02860-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-PRO-10-131-05152023

Quant Time: May 22 05:17:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

