

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051923\
 Data File : VX035765.D
 Acq On : 19 May 2023 17:04
 Operator : JC/MD
 Sample : 02860-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-PRO-05-91-0512023

Quant Time: May 22 05:18:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	197850	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	357116	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	305664	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132722	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	178578	51.738	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.480%
35) Dibromofluoromethane	5.385	113	114518	47.550	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.100%
50) Toluene-d8	8.646	98	420315	47.242	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.480%
62) 4-Bromofluorobenzene	11.079	95	159168	44.431	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.860%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.318	96	773	0.351	ug/l	84
16) Acetone	2.398	43	2436	1.642	ug/l	96
44) Trichloroethene	7.128	130	1713828	635.555	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051923\
Data File : VX035765.D
Acq On : 19 May 2023 17:04
Operator : JC/MD
Sample : 02860-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-PRO-05-91-0512023

Quant Time: May 22 05:18:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

