

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051923\
 Data File : VX035766.D
 Acq On : 19 May 2023 17:27
 Operator : JC/MD
 Sample : 02860-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-PRO-05-101-0512023

Quant Time: May 22 05:19:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	170037	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	333878	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	285278	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115077	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	173656	58.542	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	117.080%#
35) Dibromofluoromethane	5.385	113	101160	44.927	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.860%
50) Toluene-d8	8.647	98	391943	47.119	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.240%
62) 4-Bromofluorobenzene	11.085	95	141894	42.366	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.740%
Target Compounds						
					Qvalue	
16) Acetone	2.410	43	20411	16.009	ug/l	97
18) Methyl Acetate	2.709	43	1822	0.387	ug/l	# 88
25) 2-Butanone	4.592	43	49412	27.997	ug/l	98
44) Trichloroethene	7.122	130	14505	5.753	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051923\
Data File : VX035766.D
Acq On : 19 May 2023 17:27
Operator : JC/MD
Sample : 02860-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-PRO-05-101-0512023

Quant Time: May 22 05:19:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

