

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051923\
 Data File : VX035769.D
 Acq On : 19 May 2023 18:37
 Operator : JC/MD
 Sample : 02860-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-EQB-09-05122023

Quant Time: May 22 05:19:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	178582	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	348232	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	314850	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132953	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	184983	59.376	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	118.760%#
35) Dibromofluoromethane	5.385	113	116493	49.604	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.200%
50) Toluene-d8	8.647	98	420561	48.475	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.960%
62) 4-Bromofluorobenzene	11.079	95	161251	46.161	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.320%
Target Compounds						
						Qvalue
16) Acetone	2.398	43	3482	2.600	ug/l	91
29) Tetrahydrofuran	5.026	42	20749	17.810	ug/l	95
30) Chloroform	5.093	83	189940	41.547	ug/l	100
44) Trichloroethene	7.129	130	776	0.295	ug/l	98
47) Bromodichloromethane	7.824	83	4995	1.375	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051923\
Data File : VX035769.D
Acq On : 19 May 2023 18:37
Operator : JC/MD
Sample : 02860-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-EQB-09-05122023

Quant Time: May 22 05:19:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

