

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051923\
 Data File : VX035770.D
 Acq On : 19 May 2023 19:00
 Operator : JC/MD
 Sample : 02860-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-EQB-10-05152023

Quant Time: May 22 05:20:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	209282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	382084	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	331654	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	145745	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	189297	51.848	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.700%
35) Dibromofluoromethane	5.385	113	123131	47.785	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.580%
50) Toluene-d8	8.647	98	454146	47.708	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.420%
62) 4-Bromofluorobenzene	11.079	95	174698	45.579	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.160%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	3350	2.135	ug/l	99
29) Tetrahydrofuran	5.019	42	21097	15.452	ug/l	95
30) Chloroform	5.093	83	201183	37.551	ug/l	99
47) Bromodichloromethane	7.824	83	5747	1.441	ug/l	98
60) Dibromochloromethane	9.525	129	565	0.210	ug/l	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051923\
Data File : VX035770.D
Acq On : 19 May 2023 19:00
Operator : JC/MD
Sample : 02860-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-EQB-10-05152023

Quant Time: May 22 05:20:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

