

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051923\
 Data File : VX035772.D
 Acq On : 19 May 2023 19:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 22 05:20:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.652	0.674	-3.4	87	0.00
3 P	Chloromethane	0.791	0.712	10.0	81	0.00
4 C	Vinyl Chloride	0.671	0.695	-3.6#	90	0.00
5 T	Bromomethane	0.363	0.365	-0.6	90	0.00
6 T	Chloroethane	0.407	0.379	6.9	84	0.00
7 T	Trichlorofluoromethane	0.987	0.959	2.8	83	0.00
8 T	Diethyl Ether	0.380	0.356	6.3	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.586	0.586	0.0	88	0.00
10 T	Methyl Iodide	0.524	0.613	-17.0	103	0.00
11 T	Tert butyl alcohol	0.154	0.137	11.0	78	0.03
12 CM	1,1-Dichloroethene	0.557	0.542	2.7#	84	0.00
13 T	Acrolein	0.183	0.148	19.1	75	0.00
14 T	Allyl chloride	1.193	1.122	6.0	82	0.00
15 T	Acrylonitrile	0.339	0.359	-5.9	93	0.00
16 T	Acetone	0.375	0.330	12.0	81	0.00
17 T	Carbon Disulfide	1.344	1.159	13.8	73	0.00
18 T	Methyl Acetate	1.383	1.517	-9.7	96	0.00
19 T	Methyl tert-butyl Ether	2.216	2.199	0.8	88	0.00
20 T	Methylene Chloride	0.679	0.654	3.7	87	0.00
21 T	trans-1,2-Dichloroethene	0.610	0.598	2.0	84	0.00
22 T	Diisopropyl ether	2.336	2.348	-0.5	90	0.00
23 T	Vinyl Acetate	1.653	1.657	-0.2	87	0.00
24 P	1,1-Dichloroethane	1.249	1.246	0.2	87	0.00
25 T	2-Butanone	0.519	0.516	0.6	89	0.01
26 T	2,2-Dichloropropane	1.058	0.934	11.7	78	0.00
27 T	cis-1,2-Dichloroethene	0.733	0.717	2.2	86	0.00
28 T	Bromochloromethane	0.562	0.589	-4.8	103	0.00
29 T	Tetrahydrofuran	0.326	0.336	-3.1	92	0.00
30 C	Chloroform	1.280	1.282	-0.2#	88	0.00
31 T	Cyclohexane	1.102	1.091	1.0	86	0.00
32 T	1,1,1-Trichloroethane	1.114	1.103	1.0	86	0.00
33 S	1,2-Dichloroethane-d4	0.872	0.931	-6.8	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
35 S	Dibromofluoromethane	0.337	0.355	-5.3	91	0.00
36 T	1,1-Dichloropropene	0.532	0.520	2.3	85	0.00
37 T	Ethyl Acetate	0.555	0.573	-3.2	90	0.00
38 T	Carbon Tetrachloride	0.504	0.496	1.6	85	0.00
39 T	Methylcyclohexane	0.627	0.603	3.8	83	0.00
40 TM	Benzene	1.489	1.470	1.3	88	0.00
41 T	Methacrylonitrile	0.330	0.342	-3.6	92	0.00
42 TM	1,2-Dichloroethane	0.641	0.648	-1.1	90	0.00
43 T	Isopropyl Acetate	0.968	0.966	0.2	88	0.00
44 TM	Trichloroethene	0.378	0.364	3.7	86	0.00
45 C	1,2-Dichloropropane	0.394	0.394	0.0	88	0.00
46 T	Dibromomethane	0.270	0.270	0.0	87	0.00
47 T	Bromodichloromethane	0.522	0.519	0.6	86	0.00
48 T	Methyl methacrylate	0.478	0.490	-2.5	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051923\
 Data File : VX035772.D
 Acq On : 19 May 2023 19:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 22 05:20:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.008	20.0	80	0.02
50 S	Toluene-d8	1.246	1.286	-3.2	89	0.00
51 T	4-Methyl-2-Pentanone	0.558	0.600	-7.5	94	0.00
52 CM	Toluene	0.940	0.927	1.4#	88	0.00
53 T	t-1,3-Dichloropropene	0.567	0.562	0.9	83	0.00
54 T	cis-1,3-Dichloropropene	0.607	0.607	0.0	84	0.00
55 T	1,1,2-Trichloroethane	0.369	0.377	-2.2	89	0.00
56 T	Ethyl methacrylate	0.598	0.605	-1.2	88	0.00
57 T	1,3-Dichloropropane	0.669	0.674	-0.7	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.304	0.324	-6.6	90	0.00
59 T	2-Hexanone	0.420	0.444	-5.7	93	0.00
60 T	Dibromochloromethane	0.352	0.358	-1.7	86	0.00
61 T	1,2-Dibromoethane	0.389	0.393	-1.0	89	0.00
62 S	4-Bromofluorobenzene	0.502	0.530	-5.6	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.337	0.323	4.2	89	0.00
65 PM	Chlorobenzene	1.091	1.050	3.8	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.386	0.379	1.8	86	0.00
67 C	Ethyl Benzene	2.053	2.024	1.4#	88	0.00
68 T	m/p-Xylenes	0.759	0.747	1.6	87	0.00
69 T	o-Xylene	0.753	0.745	1.1	88	0.00
70 T	Styrene	1.213	1.235	-1.8	89	0.00
71 P	Bromoform	0.254	0.241	5.1	81	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	4.213	4.134	1.9	89	0.00
74 T	N-amyl acetate	1.964	1.863	5.1	88	0.00
75 P	1,1,2,2-Tetrachloroethane	1.367	1.344	1.7	92	0.00
76 T	1,2,3-Trichloropropane	1.223	1.160	5.2	92	0.00
77 T	Bromobenzene	0.956	0.911	4.7	89	0.00
78 T	n-propylbenzene	4.953	4.900	1.1	89	0.00
79 T	2-Chlorotoluene	3.101	2.996	3.4	90	0.00
80 T	1,3,5-Trimethylbenzene	3.581	3.520	1.7	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.379	0.348	8.2	81	0.00
82 T	4-Chlorotoluene	3.571	3.464	3.0	89	0.00
83 T	tert-Butylbenzene	3.481	3.358	3.5	89	0.00
84 T	1,2,4-Trimethylbenzene	3.596	3.547	1.4	90	0.00
85 T	sec-Butylbenzene	4.370	4.321	1.1	89	0.00
86 T	p-Isopropyltoluene	3.593	3.560	0.9	88	0.00
87 T	1,3-Dichlorobenzene	1.817	1.719	5.4	87	0.00
88 T	1,4-Dichlorobenzene	1.834	1.738	5.2	86	0.00
89 T	n-Butylbenzene	3.268	3.220	1.5	87	0.00
90 T	Hexachloroethane	0.545	0.526	3.5	84	0.00
91 T	1,2-Dichlorobenzene	1.779	1.705	4.2	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.304	0.297	2.3	89	0.00
93 T	1,2,4-Trichlorobenzene	1.071	0.992	7.4	83	0.00
94 T	Hexachlorobutadiene	0.425	0.386	9.2	82	0.00
95 T	Naphthalene	3.668	3.613	1.5	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051923\
Data File : VX035772.D
Acq On : 19 May 2023 19:47
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 22 05:20:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.053	0.993	5.7	85	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5