

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046256.D
 Acq On : 19 May 2025 10:39
 Operator : JC/MD
 Sample : VX0519MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0519MBL01

Quant Time: May 19 23:06:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	63875	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	126268	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121867	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52500	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	64149	53.869	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.740%	
35) Dibromofluoromethane	5.379	113	46655	51.311	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.620%	
50) Toluene-d8	8.647	98	158856	50.477	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.960%	
62) 4-Bromofluorobenzene	11.079	95	61486	50.934	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.860%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046256.D
Acq On : 19 May 2025 10:39
Operator : JC/MD
Sample : VX0519MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0519MBL01

Quant Time: May 19 23:06:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

