

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046262.D
 Acq On : 19 May 2025 13:03
 Operator : JC/MD
 Sample : Q2057-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-L

Quant Time: May 19 23:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	65399	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129752	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127866	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	55408	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67890	55.682	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	111.360%	
35) Dibromofluoromethane	5.379	113	49136	52.588	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	105.180%	
50) Toluene-d8	8.647	98	170036	52.579	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	105.160%	
62) 4-Bromofluorobenzene	11.079	95	65926	53.145	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	106.300%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	20250	41.423	ug/l	95
18) Methyl Acetate	2.703	43	1845	1.626	ug/l	95
20) Methylene Chloride	2.788	84	3800	4.056	ug/l	96
43) Isopropyl Acetate	6.342	43	11047	4.669	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046262.D
Acq On : 19 May 2025 13:03
Operator : JC/MD
Sample : Q2057-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-L

Quant Time: May 19 23:09:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

