

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046263.D
 Acq On : 19 May 2025 13:27
 Operator : JC/MD
 Sample : Q2062-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 L1-WC-1

Quant Time: May 19 23:10:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	60394	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	120720	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	116840	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49136	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62482	55.493	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.980%
35) Dibromofluoromethane	5.379	113	45097	51.877	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.760%
50) Toluene-d8	8.647	98	155498	51.681	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.360%
62) 4-Bromofluorobenzene	11.079	95	58836	50.978	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.960%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	54456	120.625	ug/l	96
18) Methyl Acetate	2.709	43	1665	1.589	ug/l	97
20) Methylene Chloride	2.782	84	9040	10.449	ug/l #	88
43) Isopropyl Acetate	6.361	43	2914	1.324	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046263.D
Acq On : 19 May 2025 13:27
Operator : JC/MD
Sample : Q2062-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
L1-WC-1

Quant Time: May 19 23:10:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

