

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046265.D
 Acq On : 19 May 2025 14:14
 Operator : JC/MD
 Sample : Q2062-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 L1-WC-3

Quant Time: May 19 23:10:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	59357	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	118499	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	114246	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49048	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60826	54.966	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.940%	
35) Dibromofluoromethane	5.385	113	43974	51.533	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.060%	
50) Toluene-d8	8.647	98	152551	51.652	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.300%	
62) 4-Bromofluorobenzene	11.079	95	58370	51.523	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.040%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	25558	57.602	ug/l	97
18) Methyl Acetate	2.709	43	2114	2.053	ug/l	96
20) Methylene Chloride	2.782	84	4105	4.828	ug/l	90
43) Isopropyl Acetate	6.349	43	7802	3.610	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046265.D
Acq On : 19 May 2025 14:14
Operator : JC/MD
Sample : Q2062-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
L1-WC-3

Quant Time: May 19 23:10:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

