

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046268.D
 Acq On : 19 May 2025 15:24
 Operator : JC/MD
 Sample : Q2062-24
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 L1-WC-6

Quant Time: May 19 23:12:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	58371	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	118055	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	113715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	45903	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59943	55.083	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.160%	
35) Dibromofluoromethane	5.385	113	43948	51.696	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.400%	
50) Toluene-d8	8.647	98	150397	51.114	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.220%	
62) 4-Bromofluorobenzene	11.079	95	55985	49.603	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 99.200%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	43583	99.886	ug/l	98
18) Methyl Acetate	2.709	43	2224	2.197	ug/l #	91
20) Methylene Chloride	2.782	84	5321	6.364	ug/l #	94
25) 2-Butanone	4.574	43	6515	10.285	ug/l #	87
40) Benzene	6.038	78	10230	3.058	ug/l	94
43) Isopropyl Acetate	6.348	43	11076	5.145	ug/l	95
52) Toluene	8.720	92	2620	1.277	ug/l	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046268.D
Acq On : 19 May 2025 15:24
Operator : JC/MD
Sample : Q2062-24
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
L1-WC-6

Quant Time: May 19 23:12:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

