

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046281.D
 Acq On : 19 May 2025 20:29
 Operator : JC/MD
 Sample : Q2073-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GDW2

Quant Time: May 19 23:19:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	61746	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	125249	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	113628	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	45883	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62475	54.272	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.540%	
35) Dibromofluoromethane	5.379	113	45215	50.132	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.260%	
50) Toluene-d8	8.647	98	155594	49.843	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.680%	
62) 4-Bromofluorobenzene	11.079	95	58373	48.748	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 97.500%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.307	50	1006	1.098	ug/l	95
16) Acetone	2.386	43	2378	5.152	ug/l	94
18) Methyl Acetate	2.709	43	1455	1.358	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046281.D
Acq On : 19 May 2025 20:29
Operator : JC/MD
Sample : Q2073-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GDW2

Quant Time: May 19 23:19:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

