

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052019\
 Data File : VX009718.D
 Acq On : 20 May 2019 12:08
 Operator : JC/SP
 Sample : K2891-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB202-EB-20190515

Quant Time: May 20 14:15:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	172339	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	268021	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	232031	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	89285	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	112613	47.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.74%	
35) Dibromofluoromethane	5.50	113	82599	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
50) Toluene-d8	8.71	98	339673	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
62) 4-Bromofluorobenzene	11.13	95	104327	42.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.42%	
Target Compounds						
16) Acetone	2.45	43	8878	7.028	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052019\
Data File : VX009718.D
Acq On : 20 May 2019 12:08
Operator : JC/SP
Sample : K2891-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB202-EB-20190515

Quant Time: May 20 14:15:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

